

Calibration Certificate

Al Takamul Yard, North Rumailah Iraq

• Phone : +964 7810009138 • www.qualitycontrol-iraq • E-mail: op@qualitycontrol-iraq

Date of Issue: September 3, 2024

Page 1 of 1

REQUEST NUMBER : By Mail	CUSTOMERS DETAILS
QC JOB NUMBER : QC/JN/24/00341	Name : Halliburton Worldwide-Iraq Branch (Sperry)
CERTIFICATE NUMBER : QC240903-31	Address : Oil Street, Western Burjessia Basra South Iraq

EQUIPMENT IDENTIFICATION AND SPECIFICATIONS

Description : Torque Wrench	Calibration Date : September 3, 2024
Type of Indication : Analog	Calibration Due : September 2, 2025
Manufacturer : PRECISION INSTRUMENTS	Last Calibration : June 7, 2023
Model : C3FR250F	
Serial Number : 276446	
SAP Number : 300189674	
Calibrated Range : 50 to 250 LB-FT	
Resolution : 5 LB-FT	
Tolerance : $\pm 4\%$	
As Found : In Tolerance	



ENVIRONMENTAL CONDITIONS DURING TEST

Ambient Temperature : 22 °C $\pm$ 2°C	Relative Humidity : 15 %RH $\pm$ 5% RH
---------------------------------------	--

CALIBRATION METHOD

The above equipment has been calibrated in accordance with International Calibration standard # ISO 6789-1:2017

The deviations of the measurements obtained from UUC with respect to reference standards are determined to obtain the error.

REFERENCE EQUIPMENT USED :

DESCRIPTION	MAKE / MODEL	SERIAL #	CALIBRATION DATE	CALIBRATION DUE DATE
Torque Wrench Calibrator	Nobar / Pro Test 1500	90962	20 Nov, 2023	19 Nov, 2024

CALIBRATION TEST RESULTS

Measurement Data For Torque

Clockwise Measurements

Readings on UUC	Readings on Ref. Standard (Avg. of 5 Measurements)	Deviation	Error Percentage (Tolerance $\pm 4\%$)	Uncertainty (95 % C.L.)
LB-FT	LB-FT	LB-FT	%	$\pm$ LB-FT
50.0	53.5	-3.50	-1.4	0.5
120.0	123.7	-3.70	-1.5	0.5
180.0	184.90	-4.90	-2.0	0.5
250.0	255.40	-5.40	-2.2	0.5

Calibration results were found to conform as per specified accuracy requirements. Above Instrument has **PASSED** its Calibration.

The reported expanded uncertainty is based on a standard uncertainty multiplied by a coverage factor $k=2$, providing a level of confidence of approximately 95%. The uncertainty evaluation has been carried out in accordance with international practice.

DEVIATION FROM STANDARD METHOD : None

REMARK (S) : ☒ The results are as found (no adjustment done).

☐ The results are post adjustment.

CALIBRATED BY	REVIEWED & APPROVED BY LAB INCHARGE	CLIENT
 Hussein Alaa	 LAB INCHARGE Asjad Rafiq	