

Certificate of Calibration

Al Takamul Yard, North Rumailah South Iraq

• Phone : +964 7810009138 • www.qualitycontrol-iraq.com • E-mail: op@qualitycontrol-iraq.com

Date of Issue: September 3, 2024

Page 1 of 1

REQUEST NUMBER	: By Mail	CUSTOMER DETAILS	
JOB NUMBER	: QC/JN/24/00341	Name	: Halliburton Worldwide IRAQ (Sperry)
CERTIFICATE NUMBER	: QC240903-27	Address	: Western Burjesia, Oil Street, District Zubair-South Iraq

EQUIPMENT IDENTIFICATION AND SPECIFICATIONS

Description : **Digimatic Micrometer**
 Manufacturer : Fowler
 Model : NA
 Serial Number : MIC-0001
 SAP No : NA
 Type : Digital
 Range : 7.00000 to 8.00000 inch
 Calibration Range : 7.00000 to 8.00000 inch
 Readability : 0.00005 inch
 As Found : In Tolerance
 Calibration Date : September 3, 2024
Calibration Due : 02 September 2025
 Last Calibration : May 30, 2023



ENVIRONMENTAL CONDITIONS DURING TEST

Ambient Temperature : 22.0 ± 2°C Humidity : 40.0 ± 5 % RH

CALIBRATION METHOD

The above equipment has been calibrated in accordance with ASME B89.1.13
 The reported uncertainty is the expanded uncertainty with k=2 according to the
 ISO/IEC "Guide to the expansion of uncertainty in measurement" providing a level of confidence of approximately 95%.

REFERENCE EQUIPMENT USED :

DESCRIPTION	SERIAL NO.	MAKE / MODEL	CALIBRATION DATE	DUE DATE
Gauge Block Set	87300	Mitutoyo / 99MAK004B2	November 15, 2023	November 14, 2024

CALIBRATION TEST RESULTS

Ref. Standard Values (A)	Measured Values on Unit Under Calibration (M)	Deviation (D=M-A)	Expanded Uncertainty (U)
inch	inch	inch	±inch
7.10000	7.10000	0.00000	0.0000
7.40000	7.40000	0.00000	0.0001
7.60000	7.60000	0.00000	0.0001
7.80000	7.80005	0.00005	0.0001
8.00000	8.00000	0.00000	0.0001

Calibration results were found to conform as per specified accuracy requirements. Above Instrument has **PASSED** its Calibration.

- ☒ The results are as found (no adjustment done).
☐ The results are post adjustment.

CALIBRATED BY	REVIEWED & APPROVED BY	LAB INCHARGE	CLIENT
 Hussein Alaa	 Asjad Rafiq	 LAB INCHARGE	