

Certificate of Calibration

Al Takamul Yard, North Rumailah South Iraq

●Phone : +964 7810009138 ● www.qualitycontrol-iraq.com ● E-mail: op@qualitycontrol-iraq.com

Date of Issue: September 6, 2024

Page 1 of 1

REQUEST NUMBER	:	By Mail	APPROVED BY LAB INCHARGE QC
JOB NUMBER	:	QC/JN/24/00343	
CERTIFICATE NUMBER	:	QC240906-03	



CUSTOMER DETAILS

Name : **Halliburton Worldwide IRAQ**
Address : Western Burjesia, Oil Street, District Zubair-South Iraq
Department : Sperry

EQUIPMENT IDENTIFICATION AND SPECIFICATIONS

Description : **GAP Gauge**
Manufacturer : N/A
Model Number : N/A
Serial Number : BAT-01
SAP No. : 300093824
Nominal Size : 1" , 0.811" , 0.04"
Resolution : NA
As Found : In Tolerance
Calibration Date : September 6, 2024
Calibration Due : **September 5, 2025**
Last Calibration Date : September 11, 2023



ENVIRONMENTAL CONDITIONS DURING TEST

Ambient Temperature : 26.0 ± 1°C Humidity : 40.0 ± 5 % RH Atm. Pressure 1013.6 ± 10 hPa

CALIBRATION METHOD

The above equipment has been calibrated in accordance with Calibration Procedure QC/CP/M/03
The deviations of the measurements obtained from UUC with respect to reference standards are determined to obtain the error.

REFERENCE EQUIPMENT USED :

DESCRIPTION	SERIAL NO.	MAKE / MODEL	CALIBRATION DATE	CALIBRATION DUE DATE
Digital Vernier Caliper	07414025	Mitutoyo / CD-12CX	16-Apr-24	15-Oct-24

CALIBRATION TEST RESULTS

Nominal Size	Reference Standard Value	Deviation	Result (Pass / Fail)
1"	1"	0.00	Pass
0.811"	0.811"	0.00	Pass
0.040"	0.040"	0.00	Pass

Calibration results were found to conform as per specified accuracy requirements. Above Instrument has **PASSED** its Calibration.

DEVIATION FROM STANDARD METHOD : None

REMARK (S) :
☒ The results are as found (no adjustment done).
☐ The results are post adjustment.

CALIBRATED BY	REVIEWED & APPROVED BY LAB INCHARGE	CLIENT
 Hussein Alaa	 LAB INCHARGE Asjad Rafiq	



Address : North Rumaila, Al Takamul Yard
Contact : +9647810009138