

Certificate of Calibration

Al Takamul Yard, North Rumailah South Iraq

• Phone : +964 7810009138 • www.qualitycontrol-iraq.com • E-mail: op@qualitycontrol-iraq.com

Date of Issue: August 3, 2024

Page 1 of 1

REQUEST NUMBER : By Contract	APPROVED BY: 
JOB NUMBER : QC/JN/24/00305	
CERTIFICATE NUMBER : QC240803-07	

CUSTOMER DETAILS

Name : **SLB Basra**
Address : North Rumaila Base south Iraq
Department : Testing

EQUIPMENT IDENTIFICATION AND SPECIFICATIONS

Description : **Pressure Gauge**
Manufacturer : **MC**
Model : **NA**
Serial Number : **868150-2**
Locatin : ESD-CD-868
Type : Analog
Range : 0 to 150 psi
Calibrated Range : 0 to 150 psi
Resolution : 5 psi
Accuarcy : 2.5
Tolerance : ±1 % of Full Scale
As found : In Tolerance
Calibrated By : Hussein Alaa
Calibration Date : August 3, 2024
Calibration Due : 02 August 2025
Last Calibration Date : NA



DEVIATION FROM STANDARD METHOD : None
REMARK (S) :

- 1 No adjustments were made prior to calibration.
- 2 Overload test was not performed.
- 3 No accessories were fitted.
- 4 Values were rounded in computation.

ENVIRONMENTAL CONDITIONS DURING TEST

Ambient Temperature : 25.0 ± 2°C Humidity : 44.0 ± 5 % RH Atm. Pressure 1011.0 ± 10 hPa

CALIBRATION METHOD

The above equipment has been calibrated in accordance with Calibration Procedure QC/CP/R/03 Rev. 01
The above instrument has been calibrated in increasing and decreasing pressure mode using a Digital Pressure Calibrator as the reference standard having traceability to Internationally recognised standards.

REFERENCE EQUIPMENT USED

DESCRIPTION	SERIAL NO.	MAKE / MODEL	CALIBRATION DATE	CALIBRATION DUE DATE
Digital Pressure Gauge	468802	AMETEK / XP2i	20 Nov, 2023	19 Nov, 2024

CALIBRATION TEST RESULTS

Measured Values on Unit Under Calibration (M)	Ref. Standard Values (A)	Deviation (D=M-A)	Error Percentage
psi	psi	psi	±
0	0	0	0.0
50	50.1	-0.10	-0.1
75	75.3	-0.30	-0.2
100	100.4	-0.40	-0.3
150	150.3	-0.30	-0.2

Calibration results were found to conform as per specified accuracy requirements. Above Instrument has **PASSED** its Calibration.

- The results are as found (no adjustment done).
 The results are post adjustment.