

Certificate of Calibration

Al Takamul Yard, North Rumailah South Iraq

• Phone : +964 7810009138 • www.qualitycontrol-iraq.com • E-mail: op@qualitycontrol-iraq.com

Date of Issue: September 19, 2024

Page 1 of 1

REQUEST NUMBER	: By Mail	APPROVED BY: [Signature] CHARGE QC
JOB NUMBER	: QC/JN/24/00307	
CERTIFICATE NUMBER	: QC240919-36	



CUSTOMER DETAILS

Name : **Aberdeen Technical Services Basra Iraq**
 Address : District Zubair-South Iraq
 Department : Workshop

EQUIPMENT IDENTIFICATION AND SPECIFICATIONS

Description : **Pressure Gauge**
 Manufacturer : CJWINTER
 Model : 11072-2-125
 Location : **Aberdeen Work Shop**
 Serial Number : D247945-001
 Type : Analog
 Range : 0 to 10000 PSI
 Calibrated Range : 0 to 10000 PSI
 Readability : 100 PSI
 Calibrated By : Hussein Alaa
 Calibration Date : September 19, 2024
Calibration Due : January 18, 2025
 Calibration Site : Aberdeen Work Shop



DEVIATION FROM STANDARD METHOD : None

REMARK (S) :

1. No adjustments were made prior to calibration.
2. Overload test was not performed.
3. No accessories were fitted.
4. Values were rounded in computation.

ENVIRONMENTAL CONDITIONS DURING TEST

Ambient Temperature : 18.0 ± 1°C Humidity : 45.0 ± 5 % RH Atm. Pressure : 1013.6 ± 10 hPa

CALIBRATION METHOD

The above equipment has been calibrated in accordance with **DKD R6-1**

The above instrument has been calibrated in increasing and decreasing pressure mode using a Digital Pressure Calibrator as the reference standard having traceability to Internationally recognised standards.

TRACEABILITY

The measurements made by Quality Control Labs are traceable to TPI Certificate #

LF-22121507

The measurements made by Quality Control Labs, realize the physical units of measurements (SI), through its state of the art calibration standards that are controlled and maintained by QC.

REFERENCE EQUIPMENT USED :

DESCRIPTION	SERIAL NO.	CERTIFICATE NO.	SOURCE OF TRACEABILITY
Digital Pressure Gauge	1505119397	LF-22121507	Liftek Company Iraq

CALIBRATION TEST RESULTS

Measured Values on Unit Under Calibration (M)	Ref. Standard Values (A)	Deviation (D=M-A)	Expanded Uncertainty (U)
PSI	PSI	PSI	± PSI
0	0	0	100
2000	2014	-14	100
4000	4015	-15	100
6000	6039	-39	100
8000	8052	-52	100
10000	10065	-65	100