


## Certificate of Calibration

Al Takamul Yard, North Rumailah South Iraq

• Phone : +964 7810009138 • www.qualitycontrol-iraq.com • E-mail: op@qualitycontrol-iraq.com

Date of Issue: September 14, 2024

Page 1 of 1

REQUEST NUMBER : By Contract	APPROVED BY: 
JOB NUMBER : QC/JN/24/00353	
CERTIFICATE NUMBER : QC240914-02	

### CUSTOMER DETAILS

Name : **slb Basra**  
Address : North Rumaila Base south Iraq  
Department : Coil Tubing

### EQUIPMENT IDENTIFICATION AND SPECIFICATIONS

Description : **Pressure Gauge**  
Manufacturer : NOV Hydra Rig  
Model : NA  
Serial Number : CAL-PG-02  
Type : Analog  
Location : Unit (MPF34215Y2064)  
Range : 0 to 200 psi  
Calibrated Range : 0 to 200 psi  
Resolution : 5 psi  
Accuracy Class : 1.0  
Tolerance : 1 % of Full Scale  
As found : In Tolerance  
Calibrated By : Hussein Alaa  
Calibration Date : September 14, 2024  
**Calibration Due : 13 September 2025**



**DEVIATION FROM STANDARD METHOD : None**

**REMARK (S) :**

1. No adjustments were made prior to calibration.
2. Overload test was not performed.
3. No accessories were fitted.
4. Values were rounded in computation.

### ENVIRONMENTAL CONDITIONS DURING TEST

Ambient Temperature : 25.0 ± 2°C Humidity : 44.0 ± 5 % RH Atm. Pressure 1011.0 ± 10 hPa

### CALIBRATION METHOD

The above equipment has been calibrated in accordance with Calibration Procedure QC/CP/R/05 Rev. 01  
The above instrument has been calibrated in increasing and decreasing pressure mode using a Digital Pressure Calibrator as the reference standard having traceability to Internationally recognised standards.

### REFERENCE EQUIPMENT USED

DESCRIPTION	SERIAL NO.	MAKE / MODEL	CALIBRATION DATE	CALIBRATION DUE DATE
Digital Pressure Gauge	468802	AMETEK / XP2i	20 Nov, 2023	19 Nov, 2024

### CALIBRATION TEST RESULTS

Measured Values on Unit Under Calibration (M)	Ref. Standard Values (A)	Deviation (D=M-A)	Error Percentage
psi	psi	psi	±
0	0.00	0.00	0.0
50	50.1	-0.10	-0.1
100	100.4	-0.40	-0.2
150	150.3	-0.30	-0.2
200	200.7	-0.70	-0.3

Calibration results were found to conform as per specified accuracy requirements. Above Instrument has **PASSED** its Calibration.

- The results are as found (no adjustment done).
- The results are post adjustment.