

Certificate of Calibration

Al Takamul Yard, North Rumailah South Iraq

• Phone : +964 7810009138 • www.qualitycontrol-iraq.com • E-mail: op@qualitycontrol-iraq.com

Date of Issue: September 17, 2024

Page 1 of 1

WORK ORDER NUMBER : By Contract JOB NUMBER : QC/JN/24/00356 CERTIFICATE NUMBER : QC240917-26	CUSTOMER DETAILS Name : Halliburton Worldwide IRAQ (TSS) Address : Western Burjesia, Oil Street, District Zubair-South Iraq
--	--

EQUIPMENT IDENTIFICATION AND SPECIFICATIONS

Description : Pressure Relief Valve
Manufacturer : BIS Valves
Model : NA
Serial Number : 13311306-09
Location : ESD PANEL (SAP # 13311306)
Set/Pressure : 10000 psi
Tolerance : ± 3%
As Found : In Tolerance
Calibration Date : September 17, 2024
Calibration Due : **16 September 2025**
Last Calibration : October 2, 2023



ENVIRONMENTAL CONDITIONS DURING TEST

Ambient Temperature : 35.0 ± 2°C **Humidity** : 30.0 ± 5 % RH **Atm. Pressure** : 1011.0 ± 10 hPa

CALIBRATION METHOD

Standard Method API 527 was used to test the unit under test. Test pressure was applied on inlet of unit under test and test was repeated 3 times. Identical repeatings were carried out as recommended in used standard.

REFERENCE EQUIPMENT USED

DESCRIPTION	SERIAL NO.	MAKE / MODEL	CALIBRATION DATE	CALIBRATION DUE DATE
Digital Pressure Gauge	1505119397	Sino Instruments / HS108	20 Nov, 2023	19 Nov, 2024

CALIBRATION TEST RESULTS

Set Pressure on Unit Under Calibration (M)	Ref. Standard Values (A)	Deviation (D=M-A)	Tolerance Limit
psi	psi	psi	± psi
10000	10016	-16	300.0

Calibration results were found to conform as per specified accuracy requirements. Above Instrument has **PASSED** its Calibration.

DEVIATION FROM STANDARD METHOD : None

REMARK (S) :

- The results are as found (no adjustment done).
- The results are post adjustment.

CALIBRATED BY  Hussein Alaa	REVIEWED & APPROVED BY  LAB INCHARGE Asjad Rafiq	CLIENT
---	--	---------------