

Certificate of Calibration

Al Takamul Yard, North Rumailah South Iraq

• Phone : +964 7810009138 • www.qualitycontrol-iraq.com • E-mail: op@qualitycontrol-iraq.com

Date of Issue: September 26, 2024

Page 1 of 1

WORK ORDER NUMBER : By Contract JOB NUMBER : QC/JN/24/00367 CERTIFICATE NUMBER : QC240926-07	CUSTOMER DETAILS Name : Halliburton Worldwide IRAQ (TSS) Address : Western Burjesia, Oil Street, District Zubair-South Iraq
---	---

EQUIPMENT IDENTIFICATION AND SPECIFICATIONS

Description : **Pressure Relief Valve**
 Manufacturer : NA
 Model : NA
 Serial Number : SPXCAL4232352
 Location : AIR COMPRESSOR
 Set/Pressure : 135 psi
 Tolerance : ± 3%
 As Found : In Tolerance
 Calibration Date : September 26, 2024
Calibration Due : 25 September 2026
 Last Calibration : December 3, 2022



ENVIRONMENTAL CONDITIONS DURING TEST

Ambient Temperature : 35.0 ± 2°C Humidity : 30.0 ± 5 % RH Atm. Pressure : 1011.0 ± 10 hPa

CALIBRATION METHOD

Standard Method API 527 was used to test the unit under test. Test pressure was applied on inlet of unit under test and test was repeated 3 times. Identical repeatings were carried out as recommended in used standard.

REFERENCE EQUIPMENT USED

DESCRIPTION	SERIAL NO.	MAKE / MODEL	CALIBRATION DATE	CALIBRATION DUE DATE
Digital Pressure Gauge	1505119397	Sino Instruments / HS108	20 Nov, 2023	19 Nov, 2024

CALIBRATION TEST RESULTS

Set Pressure on Unit Under Calibration (M)	Ref. Standard Values (A)	Deviation (D=M-A)	Tolerance Limit
psi	psi	psi	± psi
135	136.3	-1.3	4.1

Calibration results were found to conform as per specified accuracy requirements. Above Instrument has **PASSED** its Calibration.

DEVIATION FROM STANDARD METHOD : None

REMARK (S) :

- The results are as found (no adjustment done).
- The results are post adjustment.

CALIBRATED BY  Hussein Alaa	REVIEWED & APPROVED BY  LAB INCHARGE Asjad Rafiq	CLIENT
--	--	--------