

Certificate of Calibration

Al Takamul Yard, North Rumailah South Iraq

• Phone : +964 7810009138 • www.qualitycontrol-iraq.com • E-mail: op@qualitycontrol-iraq.com

Date of Issue: September 20, 2024

Page 1 of 1

REQUEST NUMBER : By Contract	CUSTOMER DETAILS
JOB NUMBER : QC/JN/24/00368	Name Halliburton Worldwide IRAQ (TSS)
CERTIFICATE NUMBER : QC240920-04	Address Western Burjesia, Oil Street, District Zubair-South Iraq

EQUIPMENT IDENTIFICATION AND SPECIFICATIONS

Description : **Pressure Gauge**
 Manufacturer : WIKA
 Model : EN837-1
 Serial Number : 2302211104-4
 Type : Analog
 Location : SPARE
 Range : 0 to 300 psi
 Calibrated Range : 0 to 300 psi
 Resolution : 5 psi
 Accuracy : 1.0
 Tolerance : ± 1.0 % of Full Scale
 As found : In Tolerance
 Calibration Date : September 20, 2024
Calibration Due : 19 September 2025
 Last Calibration Date : NA



DEVIATION FROM STANDARD METHOD : None

REMARK (S) :

1. No adjustments were made prior to calibration.
2. No accessories were fitted.
3. Values were rounded in computation.

ENVIRONMENTAL CONDITIONS DURING TEST

Ambient Temperature : 22.0 ± 2°C Humidity : 15.0 ± 5 % RH Atm. Pressure : 1012.0 ± 10 hPa

CALIBRATION METHOD

The above equipment has been calibrated in accordance with Calibration Procedure QC/CP/R/03 Rev. 01

The above instrument has been calibrated in increasing and decreasing pressure mode using a Digital Pressure Calibrator as the reference standard having traceability to Internationally recognised standards.

REFERENCE EQUIPMENT USED

DESCRIPTION	SERIAL NO.	MAKE / MODEL	CALIBRATION DATE	CALIBRATION DUE DATE
Digital Pressure Gauge	468802	AMETEK / XP2i	20 Nov, 2023	19 Nov, 2024

CALIBRATION TEST RESULTS

Measured Values on Unit Under Calibration (M)	Ref. Standard Values (A)	Deviation (D=M-A)	Tolerance Limit
psi	psi	psi	± psi
0	0.0	0.0	0.0
75	75.3	-0.3	-0.1
150	150.5	-0.5	-0.2
225	225.6	-0.6	-0.2
300	300.8	-0.8	-0.3

Calibration results were found to conform as per specified accuracy requirements. Above Instrument has **PASSED** its Calibration.

- The results are as found (no adjustment done).
 The results are post adjustment.

CALIBRATED BY  Hussein Alaa	REVIEWED & APPROVED BY LAB INCHARGE  LAB INCHARGE Asjad Rafiq	CLIENT
--	---	--------

