

Certificate of Calibration

Al Takamul Yard, North Rumailah South Iraq

• Phone : +964 7810009138 • www.qualitycontrol-iraq.com • E-mail: op@qualitycontrol-iraq.com

Date of Issue: October 15, 2024

Page 1 of 1

REQUEST NUMBER : By Contract	CUSTOMER DETAILS
JOB NUMBER : QC/JN/24/00385	Name Halliburton Worldwide IRAQ (HPS)
CERTIFICATE NUMBER : QC241015-27	Address Western Burjesia, Oil Street, District Zubair-South Iraq

Description : **Pressure Gauge**
 Manufacturer : NOSHOK
 Model : NA
 Serial Number : **CAL-PG-27**
 Type : Analog
 Location : Power Pack (SAP 13410135)
 Range : 0 to 6000 psi
 Calibrated Range : 0 to 6000 psi
 Resolution : 100 psi
 Accuracy : NA
 Tolerance : ± 1.0 % of Full Scale
 As found : In Tolerance
 Calibration Date : October 15, 2024
Calibration Due : 14 October 2025
 Last Calibration : October 12, 2023



DEVIATION FROM STANDARD METHOD : None

REMARK (S) :

1. No adjustments were made prior to calibration.
2. No accessories were fitted.
3. Values were rounded in computation.

ENVIRONMENTAL CONDITIONS DURING TEST

Ambient Temperature : 35.0 ± 2°C Humidity : 15.0 ± 5 % RH Atm. Pressure : 1013.0 ± 10 hPa

CALIBRATION METHOD

The above equipment has been calibrated in accordance with Calibration Procedure QC/CP/R/03 Rev. 01

The above instrument has been calibrated in increasing and decreasing pressure mode using a Digital Pressure Calibrator as the reference standard having traceability to Internationally recognised standards.

REFERENCE EQUIPMENT USED


DESCRIPTION	SERIAL NO.	MAKE / MODEL	CALIBRATION DATE	CALIBRATION DUE DATE
Digital Pressure Gauge	468802	AMETEK / XP2i	20 Nov, 2023	19 Nov, 2024

CALIBRATION TEST RESULTS

Measured Values on Unit Under Calibration (M)	Ref. Standard Values (A)	Deviation (D=M-A)	Full Scale Error %
psi	psi	psi	%
0	0.00	0.00	0.0
1500	1511.0	-11.00	-0.2
3000	3013.0	-13.00	-0.2
4500	4517.0	-17.00	-0.3
6000	6020.0	-20.00	-0.3

Calibration results were found to conform as per specified accuracy requirements. Above Instrument has **PASSED** its Calibration.

- The results are as found (no adjustment done).
 The results are post adjustment.

CALIBRATED BY	REVIEWED & APPROVED BY LAB INCHARGE	CLIENT
 Abdulrahman Loay	 LAB INCHARGE Asjad Rafiq	

Certificate of Calibration

Al Takamul Yard, North Rumailah South Iraq

• Phone : +964 7810009138 • www.qualitycontrol-iraq.com • E-mail: op@qualitycontrol-iraq.com

Date of Issue: October 15, 2024

Page 1 of 1

REQUEST NUMBER : By Contract	CUSTOMER DETAILS
JOB NUMBER : QC/JN/24/00385	Name Halliburton Worldwide IRAQ (HPS)
CERTIFICATE NUMBER : QC241015-28	Address Western Burjesia, Oil Street, District Zubair-South Iraq

Description : **Pressure Gauge**
 Manufacturer : NOSHOK
 Model : NA
 Serial Number : **CAL-PG-28**
 Type : Analog
 Location : Power Pack (SAP 13410135)
 Range : 0 to 5000 psi
 Calibrated Range : 0 to 5000 psi
 Resolution : 100 psi
 Accuracy : NA
 Tolerance : ± 1.0 % of Full Scale
 As found : In Tolerance
 Calibration Date : October 15, 2024
Calibration Due : 14 October 2025
 Last Calibration : October 12, 2023



DEVIATION FROM STANDARD METHOD : None

REMARK (S) :

1. No adjustments were made prior to calibration.
2. No accessories were fitted.
3. Values were rounded in computation.

ENVIRONMENTAL CONDITIONS DURING TEST

Ambient Temperature : 35.0 ± 2°C Humidity : 15.0 ± 5 % RH Atm. Pressure : 1013.0 ± 10 hPa

CALIBRATION METHOD

The above equipment has been calibrated in accordance with Calibration Procedure QC/CP/R/03 Rev. 01

The above instrument has been calibrated in increasing and decreasing pressure mode using a Digital Pressure Calibrator as the reference standard having traceability to Internationally recognised standards.

REFERENCE EQUIPMENT USED



DESCRIPTION	SERIAL NO.	MAKE / MODEL	CALIBRATION DATE	CALIBRATION DUE DATE
Digital Pressure Gauge	468802	AMETEK / XP2i	20 Nov, 2023	19 Nov, 2024

CALIBRATION TEST RESULTS

Measured Values on Unit Under Calibration (M)	Ref. Standard Values (A)	Deviation (D=M-A)	Full Scale Error %
psi	psi	psi	%
0	0.00	0.00	0.0
1250	1254.0	-4.00	-0.1
2500	2505.0	-5.00	-0.1
3750	3758.0	-8.00	-0.2
5000	5008.0	-8.00	-0.2

Calibration results were found to conform as per specified accuracy requirements. Above Instrument has **PASSED** its Calibration.

- The results are as found (no adjustment done).
 The results are post adjustment.

CALIBRATED BY	REVIEWED & APPROVED BY LAB INCHARGE	CLIENT
 Abdulrahman Loay	 LAB INCHARGE Asjad Rafiq	

Certificate of Calibration

Al Takamul Yard, North Rumailah South Iraq

• Phone : +964 7810009138 • www.qualitycontrol-iraq.com • E-mail: op@qualitycontrol-iraq.com

Date of Issue: October 15, 2024

Page 1 of 1

REQUEST NUMBER : By Contract	CUSTOMER DETAILS
JOB NUMBER : QC/JN/24/00385	Name Halliburton Worldwide IRAQ (HPS)
CERTIFICATE NUMBER : QC241015-29	Address Western Burjesia, Oil Street, District Zubair-South Iraq

Description : **Pressure Gauge**
 Manufacturer : NOSHOK
 Model : NA
 Serial Number : **CAL-PG-29**
 Type : Analog
 Location : Power Pack (SAP 13410135)
 Range : 0 to 3000 psi
 Calibrated Range : 0 to 3000 psi
 Resolution : 50 psi
 Accuracy : NA
 Tolerance : ± 1.0 % of Full Scale
 As found : In Tolerance
 Calibration Date : October 15, 2024
Calibration Due : 14 October 2025
 Last Calibration : October 12, 2023



DEVIATION FROM STANDARD METHOD : None

REMARK (S) :

1. No adjustments were made prior to calibration.
2. No accessories were fitted.
3. Values were rounded in computation.

ENVIRONMENTAL CONDITIONS DURING TEST

Ambient Temperature : 35.0 ± 2°C Humidity : 15.0 ± 5 % RH Atm. Pressure : 1013.0 ± 10 hPa

CALIBRATION METHOD

The above equipment has been calibrated in accordance with Calibration Procedure QC/CP/R/03 Rev. 01

The above instrument has been calibrated in increasing and decreasing pressure mode using a Digital Pressure Calibrator as the reference standard having traceability to Internationally recognised standards.

REFERENCE EQUIPMENT USED



DESCRIPTION	SERIAL NO.	MAKE / MODEL	CALIBRATION DATE	CALIBRATION DUE DATE
Digital Pressure Gauge	468802	AMETEK / XP2i	20 Nov, 2023	19 Nov, 2024

CALIBRATION TEST RESULTS

Measured Values on Unit Under Calibration (M)	Ref. Standard Values (A)	Deviation (D=M-A)	Full Scale Error %
psi	psi	psi	%
0	0.00	0.00	0.0
750	758.0	-8.00	-0.3
1500	1511.0	-11.00	-0.4
2250	2263.0	-13.00	-0.4
3000	3015.0	-15.00	-0.5

Calibration results were found to conform as per specified accuracy requirements. Above Instrument has **PASSED** its Calibration.

- The results are as found (no adjustment done).
 The results are post adjustment.

CALIBRATED BY	REVIEWED & APPROVED BY LAB INCHARGE	CLIENT
 Abdulrahman Loay	 LAB INCHARGE Asjad Rafiq	

Certificate of Calibration

Al Takamul Yard, North Rumailah South Iraq

• Phone : +964 7810009138 • www.qualitycontrol-iraq.com • E-mail: op@qualitycontrol-iraq.com

Date of Issue: October 15, 2024

Page 1 of 1

REQUEST NUMBER : By Contract	CUSTOMER DETAILS
JOB NUMBER : QC/JN/24/00385	Name Halliburton Worldwide IRAQ (HPS)
CERTIFICATE NUMBER : QC241015-30	Address Western Burjesia, Oil Street, District Zubair-South Iraq

Description : **Pressure Gauge**
 Manufacturer : NOSHOK
 Model : NA
 Serial Number : **CAL-PG-30**
 Type : Analog
 Location : Power Pack (SAP 13410135)
 Range : 0 to 3000 psi
 Calibrated Range : 0 to 3000 psi
 Resolution : 50 psi
 Accuracy : NA
 Tolerance : ± 1.0 % of Full Scale
 As found : In Tolerance
 Calibration Date : October 15, 2024
Calibration Due : 14 October 2025
 Last Calibration : October 12, 2023



DEVIATION FROM STANDARD METHOD : None

REMARK (S) :

1. No adjustments were made prior to calibration.
2. No accessories were fitted.
3. Values were rounded in computation.

ENVIRONMENTAL CONDITIONS DURING TEST

Ambient Temperature : 35.0 ± 2°C Humidity : 15.0 ± 5 % RH Atm. Pressure : 1013.0 ± 10 hPa

CALIBRATION METHOD

The above equipment has been calibrated in accordance with Calibration Procedure QC/CP/R/03 Rev. 01

The above instrument has been calibrated in increasing and decreasing pressure mode using a Digital Pressure Calibrator as the reference standard having traceability to Internationally recognised standards.

REFERENCE EQUIPMENT USED



DESCRIPTION	SERIAL NO.	MAKE / MODEL	CALIBRATION DATE	CALIBRATION DUE DATE
Digital Pressure Gauge	468802	AMETEK / XP2i	20 Nov, 2023	19 Nov, 2024

CALIBRATION TEST RESULTS

Measured Values on Unit Under Calibration (M)	Ref. Standard Values (A)	Deviation (D=M-A)	Full Scale Error %
psi	psi	psi	%
0	0.00	0.00	0.0
750	735.0	15.00	0.5
1500	1487.0	13.00	0.4
2250	2238.0	12.00	0.4
3000	2990.0	10.00	0.3

Calibration results were found to conform as per specified accuracy requirements. Above Instrument has **PASSED** its Calibration.

- The results are as found (no adjustment done).
 The results are post adjustment.

CALIBRATED BY	REVIEWED & APPROVED BY LAB INCHARGE	CLIENT
 Abdulrahman Loay	 LAB INCHARGE Asjad Rafiq	

Certificate of Calibration

Al Takamul Yard, North Rumailah South Iraq

• Phone : +964 7810009138 • www.qualitycontrol-iraq.com • E-mail: op@qualitycontrol-iraq.com

Date of Issue: October 15, 2024

Page 1 of 1

REQUEST NUMBER : By Contract	CUSTOMER DETAILS
JOB NUMBER : QC/JN/24/00385	Name Halliburton Worldwide IRAQ (HPS)
CERTIFICATE NUMBER : QC241015-31	Address Western Burjesia, Oil Street, District Zubair-South Iraq

Description : **Pressure Gauge**
 Manufacturer : NOSHOK
 Model : NA
 Serial Number : **CAL-PG-31**
 Type : Analog
 Location : Power Pack (SAP 13410135)
 Range : 0 to 5000 psi
 Calibrated Range : 0 to 5000 psi
 Resolution : 100 psi
 Accuracy : NA
 Tolerance : ± 1.0 % of Full Scale
 As found : In Tolerance
 Calibration Date : October 15, 2024
Calibration Due : 14 October 2025
 Last Calibration : October 12, 2023



DEVIATION FROM STANDARD METHOD : None

REMARK (S) :

1. No adjustments were made prior to calibration.
2. No accessories were fitted.
3. Values were rounded in computation.

ENVIRONMENTAL CONDITIONS DURING TEST

Ambient Temperature : 35.0 ± 2°C Humidity : 15.0 ± 5 % RH Atm. Pressure : 1013.0 ± 10 hPa

CALIBRATION METHOD

The above equipment has been calibrated in accordance with Calibration Procedure QC/CP/R/03 Rev. 01

The above instrument has been calibrated in increasing and decreasing pressure mode using a Digital Pressure Calibrator as the reference standard having traceability to Internationally recognised standards.

REFERENCE EQUIPMENT USED



DESCRIPTION	SERIAL NO.	MAKE / MODEL	CALIBRATION DATE	CALIBRATION DUE DATE
Digital Pressure Gauge	468802	AMETEK / XP2i	20 Nov, 2023	19 Nov, 2024

CALIBRATION TEST RESULTS

Measured Values on Unit Under Calibration (M)	Ref. Standard Values (A)	Deviation (D=M-A)	Full Scale Error %
psi	psi	psi	%
0	0.00	0.00	0.0
1200	1217.0	-17.00	-0.3
2500	2520.0	-20.00	-0.4
3700	3723.0	-23.00	-0.5
5000	5025.0	-25.00	-0.5

Calibration results were found to conform as per specified accuracy requirements. Above Instrument has **PASSED** its Calibration.

- The results are as found (no adjustment done).
 The results are post adjustment.

CALIBRATED BY	REVIEWED & APPROVED BY LAB INCHARGE	CLIENT
 Abdulrahman Loay	 LAB INCHARGE Asjad Rafiq	

Certificate of Calibration

Al Takamul Yard, North Rumailah South Iraq

• Phone : +964 7810009138 • www.qualitycontrol-iraq.com • E-mail: op@qualitycontrol-iraq.com

Date of Issue: October 15, 2024

Page 1 of 1

REQUEST NUMBER : By Contract	CUSTOMER DETAILS
JOB NUMBER : QC/JN/24/00385	Name Halliburton Worldwide IRAQ (HPS)
CERTIFICATE NUMBER : QC241015-32	Address Western Burjesia, Oil Street, District Zubair-South Iraq

Description : **Pressure Gauge**
 Manufacturer : NOSHOK
 Model : NA
 Serial Number : **CAL-PG-32**
 Type : Analog
 Location : Power Pack (SAP 13410135)
 Range : 0 to 6000 psi
 Calibrated Range : 0 to 6000 psi
 Resolution : 100 psi
 Accuracy : NA
 Tolerance : ± 1.0 % of Full Scale
 As found : In Tolerance
 Calibration Date : October 15, 2024
Calibration Due : 14 October 2025
 Last Calibration : October 12, 2023



DEVIATION FROM STANDARD METHOD : None

REMARK (S) :

1. No adjustments were made prior to calibration.
2. No accessories were fitted.
3. Values were rounded in computation.

ENVIRONMENTAL CONDITIONS DURING TEST

Ambient Temperature : 35.0 ± 2°C Humidity : 15.0 ± 5 % RH Atm. Pressure : 1013.0 ± 10 hPa

CALIBRATION METHOD

The above equipment has been calibrated in accordance with Calibration Procedure QC/CP/R/03 Rev. 01
 The above instrument has been calibrated in increasing and decreasing pressure mode using a Digital Pressure Calibrator as the reference standard having traceability to Internationally recognised standards.

REFERENCE EQUIPMENT USED

DESCRIPTION	SERIAL NO.	MAKE / MODEL	CALIBRATION DATE	CALIBRATION DUE DATE
Digital Pressure Gauge	468802	AMETEK / XP2i	20 Nov, 2023	19 Nov, 2024

CALIBRATION TEST RESULTS

Measured Values on Unit Under Calibration (M)	Ref. Standard Values (A)	Deviation (D=M-A)	Full Scale Error %
psi	psi	psi	%
0	0.00	0.00	0.0
1500	1511.0	-11.00	-0.2
3000	3014.0	-14.00	-0.2
4500	4515.0	-15.00	-0.3
6000	6019.0	-19.00	-0.3

Calibration results were found to conform as per specified accuracy requirements. Above Instrument has **PASSED** its Calibration.

- The results are as found (no adjustment done).
 The results are post adjustment.

CALIBRATED BY	REVIEWED & APPROVED BY LAB INCHARGE	CLIENT
 Abdulrahman Loay	 LAB INCHARGE Asjad Rafiq	

Certificate of Calibration

Al Takamul Yard, North Rumailah South Iraq

• Phone : +964 7810009138 • www.qualitycontrol-iraq.com • E-mail: op@qualitycontrol-iraq.com

Date of Issue: October 15, 2024

Page 1 of 1

REQUEST NUMBER : By Contract	CUSTOMER DETAILS
JOB NUMBER : QC/JN/24/00385	Name Halliburton Worldwide IRAQ (HPS)
CERTIFICATE NUMBER : QC241015-33	Address Western Burjesia, Oil Street, District Zubair-South Iraq

Description : **Pressure Gauge**
 Manufacturer : NOSHOK
 Model : NA
 Serial Number : **CAL-PG-33**
 Type : Analog
 Location : Power Pack (SAP 13410135)
 Range : 0 to 5000 psi
 Calibrated Range : 0 to 5000 psi
 Resolution : 100 psi
 Accuracy : NA
 Tolerance : ± 1.0 % of Full Scale
 As found : In Tolerance
 Calibration Date : October 15, 2024
Calibration Due : 14 October 2025
 Last Calibration : October 12, 2023



DEVIATION FROM STANDARD METHOD : None

REMARK (S) :

1. No adjustments were made prior to calibration.
2. No accessories were fitted.
3. Values were rounded in computation.

ENVIRONMENTAL CONDITIONS DURING TEST

Ambient Temperature : 35.0 ± 2°C Humidity : 15.0 ± 5 % RH Atm. Pressure : 1013.0 ± 10 hPa

CALIBRATION METHOD

The above equipment has been calibrated in accordance with Calibration Procedure QC/CP/R/03 Rev. 01

The above instrument has been calibrated in increasing and decreasing pressure mode using a Digital Pressure Calibrator as the reference standard having traceability to Internationally recognised standards.

REFERENCE EQUIPMENT USED


DESCRIPTION	SERIAL NO.	MAKE / MODEL	CALIBRATION DATE	CALIBRATION DUE DATE
Digital Pressure Gauge	468802	AMETEK / XP2i	20 Nov, 2023	19 Nov, 2024

CALIBRATION TEST RESULTS

Measured Values on Unit Under Calibration (M)	Ref. Standard Values (A)	Deviation (D=M-A)	Full Scale Error %
psi	psi	psi	%
0	0.00	0.00	0.0
1250	1255.0	-5.00	-0.1
2500	2508.0	-8.00	-0.2
3750	3760.0	-10.00	-0.2
5000	5012.0	-12.00	-0.2

Calibration results were found to conform as per specified accuracy requirements. Above Instrument has **PASSED** its Calibration.

- The results are as found (no adjustment done).
 The results are post adjustment.

CALIBRATED BY	REVIEWED & APPROVED BY LAB INCHARGE	CLIENT
 Abdulrahman Loay	 LAB INCHARGE Asjad Rafiq	

Certificate of Calibration

Al Takamul Yard, North Rumailah South Iraq

• Phone : +964 7810009138 • www.qualitycontrol-iraq.com • E-mail: op@qualitycontrol-iraq.com

Date of Issue: October 15, 2024

Page 1 of 1

REQUEST NUMBER : By Contract	CUSTOMER DETAILS
JOB NUMBER : QC/JN/24/00385	Name Halliburton Worldwide IRAQ (HPS)
CERTIFICATE NUMBER : QC241015-34	Address Western Burjesia, Oil Street, District Zubair-South Iraq

Description : **Pressure Gauge**
 Manufacturer : NOSHOK
 Model : NA
 Serial Number : **CAL-PG-34**
 Type : Analog
 Location : Power Pack (SAP 13410135)
 Range : 0 to 3000 psi
 Calibrated Range : 0 to 3000 psi
 Resolution : 50 psi
 Accuracy : NA
 Tolerance : ± 1.0 % of Full Scale
 As found : In Tolerance
 Calibration Date : October 15, 2024
Calibration Due : 14 October 2025
 Last Calibration : October 12, 2023



DEVIATION FROM STANDARD METHOD : None

REMARK (S) :

1. No adjustments were made prior to calibration.
2. No accessories were fitted.
3. Values were rounded in computation.

ENVIRONMENTAL CONDITIONS DURING TEST

Ambient Temperature : 35.0 ± 2°C Humidity : 15.0 ± 5 % RH Atm. Pressure : 1013.0 ± 10 hPa

CALIBRATION METHOD

The above equipment has been calibrated in accordance with Calibration Procedure QC/CP/R/03 Rev. 01

The above instrument has been calibrated in increasing and decreasing pressure mode using a Digital Pressure Calibrator as the reference standard having traceability to Internationally recognised standards.

REFERENCE EQUIPMENT USED



DESCRIPTION	SERIAL NO.	MAKE / MODEL	CALIBRATION DATE	CALIBRATION DUE DATE
Digital Pressure Gauge	468802	AMETEK / XP2i	20 Nov, 2023	19 Nov, 2024

CALIBRATION TEST RESULTS

Measured Values on Unit Under Calibration (M)	Ref. Standard Values (A)	Deviation (D=M-A)	Full Scale Error %
psi	psi	psi	%
0	0.00	0.00	0.0
750	753.0	-3.00	-0.1
1500	1505.0	-5.00	-0.2
2250	2256.0	-6.00	-0.2
3000	3007.0	-7.00	-0.2

Calibration results were found to conform as per specified accuracy requirements. Above Instrument has **PASSED** its Calibration.

- The results are as found (no adjustment done).
 The results are post adjustment.

CALIBRATED BY	REVIEWED & APPROVED BY LAB INCHARGE	CLIENT
 Abdulrahman Loay	 LAB INCHARGE Asjad Rafiq	

Certificate of Calibration

Al Takamul Yard, North Rumailah South Iraq

• Phone : +964 7810009138 • www.qualitycontrol-iraq.com • E-mail: op@qualitycontrol-iraq.com

Date of Issue: October 15, 2024

Page 1 of 1

REQUEST NUMBER : By Contract	CUSTOMER DETAILS
JOB NUMBER : QC/JN/24/00385	Name Halliburton Worldwide IRAQ (HPS)
CERTIFICATE NUMBER : QC241015-35	Address Western Burjesia, Oil Street, District Zubair-South Iraq

Description : **Pressure Gauge**
 Manufacturer : NOSHOK
 Model : NA
 Serial Number : **CAL-PG-35**
 Type : Analog
 Location : Power Pack (SAP 13410135)
 Range : 0 to 1000 psi
 Calibrated Range : 0 to 1000 psi
 Resolution : 20 psi
 Accuracy : NA
 Tolerance : ± 1.0 % of Full Scale
 As found : In Tolerance
 Calibration Date : October 15, 2024
Calibration Due : 14 October 2025
 Last Calibration : October 12, 2023



DEVIATION FROM STANDARD METHOD : None

REMARK (S) :

1. No adjustments were made prior to calibration.
2. No accessories were fitted.
3. Values were rounded in computation.

ENVIRONMENTAL CONDITIONS DURING TEST

Ambient Temperature : 35.0 ± 2°C Humidity : 15.0 ± 5 % RH Atm. Pressure : 1013.0 ± 10 hPa

CALIBRATION METHOD

The above equipment has been calibrated in accordance with Calibration Procedure QC/CP/R/03 Rev. 01

The above instrument has been calibrated in increasing and decreasing pressure mode using a Digital Pressure Calibrator as the reference standard having traceability to Internationally recognised standards.

REFERENCE EQUIPMENT USED

DESCRIPTION	SERIAL NO.	MAKE / MODEL	CALIBRATION DATE	CALIBRATION DUE DATE
Digital Pressure Gauge	468802	AMETEK / XP2i	20 Nov, 2023	19 Nov, 2024

CALIBRATION TEST RESULTS

Measured Values on Unit Under Calibration (M)	Ref. Standard Values (A)	Deviation (D=M-A)	Full Scale Error %
psi	psi	psi	%
0	0.00	0.00	0.0
240	242.0	-2.00	-0.2
500	503.0	-3.00	-0.3
740	745.0	-5.00	-0.5
1000	1006.0	-6.00	-0.6

Calibration results were found to conform as per specified accuracy requirements. Above Instrument has **PASSED** its Calibration.

- The results are as found (no adjustment done).
 The results are post adjustment.

CALIBRATED BY	REVIEWED & APPROVED BY LAB INCHARGE	CLIENT
 Abdulrahman Loay	 LAB INCHARGE Asjad Rafiq	

Certificate of Calibration

Al Takamul Yard, North Rumailah South Iraq

• Phone : +964 7810009138 • www.qualitycontrol-iraq.com • E-mail: op@qualitycontrol-iraq.com

Date of Issue: October 15, 2024

Page 1 of 1

REQUEST NUMBER : By Contract	CUSTOMER DETAILS
JOB NUMBER : QC/JN/24/00385	Name Halliburton Worldwide IRAQ (HPS)
CERTIFICATE NUMBER : QC241015-36	Address Western Burjesia, Oil Street, District Zubair-South Iraq

Description : **Pressure Gauge**
 Manufacturer : NOSHOK
 Model : NA
 Serial Number : **CAL-PG-36**
 Type : Analog
 Location : Power Pack (SAP 13410135)
 Range : 0 to 100 psi
 Calibrated Range : 0 to 100 psi
 Resolution : 2 psi
 Accuracy : NA
 Tolerance : ± 1.0 % of Full Scale
 As found : In Tolerance
 Calibration Date : October 15, 2024
Calibration Due : 14 October 2025
 Last Calibration : October 12, 2023



DEVIATION FROM STANDARD METHOD : None

REMARK (S) :

1. No adjustments were made prior to calibration.
2. No accessories were fitted.
3. Values were rounded in computation.

ENVIRONMENTAL CONDITIONS DURING TEST

Ambient Temperature : 35.0 ± 2°C Humidity : 15.0 ± 5 % RH Atm. Pressure : 1013.0 ± 10 hPa

CALIBRATION METHOD

The above equipment has been calibrated in accordance with Calibration Procedure QC/CP/R/03 Rev. 01

The above instrument has been calibrated in increasing and decreasing pressure mode using a Digital Pressure Calibrator as the reference standard having traceability to Internationally recognised standards.

REFERENCE EQUIPMENT USED



DESCRIPTION	SERIAL NO.	MAKE / MODEL	CALIBRATION DATE	CALIBRATION DUE DATE
Digital Pressure Gauge	468802	AMETEK / XP2i	20 Nov, 2023	19 Nov, 2024

CALIBRATION TEST RESULTS

Measured Values on Unit Under Calibration (M)	Ref. Standard Values (A)	Deviation (D=M-A)	Full Scale Error %
psi	psi	psi	%
0	0.00	0.00	0.0
30	30.3	-0.30	-0.3
50	50.5	-0.50	-0.5
70	70.5	-0.50	-0.5
100	100.7	-0.70	-0.7

Calibration results were found to conform as per specified accuracy requirements. Above Instrument has **PASSED** its Calibration.

- The results are as found (no adjustment done).
 The results are post adjustment.

CALIBRATED BY	REVIEWED & APPROVED BY LAB INCHARGE	CLIENT
 Abdulrahman Loay	 LAB INCHARGE Asjad Rafiq	

Certificate of Calibration

Al Takamul Yard, North Rumailah South Iraq

• Phone : +964 7810009138 • www.qualitycontrol-iraq.com • E-mail: op@qualitycontrol-iraq.com

Date of Issue: October 15, 2024

Page 1 of 1

REQUEST NUMBER : By Contract	CUSTOMER DETAILS
JOB NUMBER : QC/JN/24/00385	Name Halliburton Worldwide IRAQ (HPS)
CERTIFICATE NUMBER : QC241015-37	Address Western Burjesia, Oil Street, District Zubair-South Iraq

Description : **Pressure Gauge**
 Manufacturer : NOV Hydra Rig
 Model : NA
 Serial Number : **CAL-PG-37**
 Type : Analog
 Location : Power Pack (SAP 13410135)
 Range : 0 to 3000 psi
 Calibrated Range : 0 to 3000 psi
 Resolution : 50 psi
 Accuracy : NA
 Tolerance : ± 1.0 % of Full Scale
 As found : In Tolerance
 Calibration Date : October 15, 2024
Calibration Due : 14 October 2025
 Last Calibration : October 12, 2023



DEVIATION FROM STANDARD METHOD : None

REMARK (S) :

1. No adjustments were made prior to calibration.
2. No accessories were fitted.
3. Values were rounded in computation.

ENVIRONMENTAL CONDITIONS DURING TEST

Ambient Temperature : 35.0 ± 2°C Humidity : 15.0 ± 5 % RH Atm. Pressure : 1013.0 ± 10 hPa

CALIBRATION METHOD

The above equipment has been calibrated in accordance with Calibration Procedure QC/CP/R/03 Rev. 01

The above instrument has been calibrated in increasing and decreasing pressure mode using a Digital Pressure Calibrator as the reference standard having traceability to Internationally recognised standards.

REFERENCE EQUIPMENT USED

DESCRIPTION	SERIAL NO.	MAKE / MODEL	CALIBRATION DATE	CALIBRATION DUE DATE
Digital Pressure Gauge	468802	AMETEK / XP2i	20 Nov, 2023	19 Nov, 2024

CALIBRATION TEST RESULTS

Measured Values on Unit Under Calibration (M)	Ref. Standard Values (A)	Deviation (D=M-A)	Full Scale Error %
psi	psi	psi	%
0	0.00	0.00	0.0
750	755.0	-5.00	-0.2
1500	1507.0	-7.00	-0.2
2250	2258.0	-8.00	-0.3
3000	3011.0	-11.00	-0.4

Calibration results were found to conform as per specified accuracy requirements. Above Instrument has **PASSED** its Calibration.

- The results are as found (no adjustment done).
 The results are post adjustment.

CALIBRATED BY	REVIEWED & APPROVED BY LAB INCHARGE	CLIENT
 Abdulrahman Loay	 LAB INCHARGE Asjad Rafiq	