

Certificate of Calibration

Al Takamul Yard, North Rumailah South Iraq

• Phone : +964 7810009138 • www.qualitycontrol-iraq.com • E-mail: op@qualitycontrol-iraq.com

Date of Issue: November 3, 2024

Page 1 of 1

REQUEST NUMBER	: By mail	APPROVED BY	AL TAKAMUL COMPANY QC
JOB NUMBER	: QC-YB-0022		
CERTIFICATE NUMBER	: QC-YB-0022-02		



CUSTOMER DETAILS

Name : Egyptian Maintenance Company(EMC) IRAQ
Address : Iraq, Basra, North Rumaila

EQUIPMENT IDENTIFICATION AND SPECIFICATIONS

Description : Pressure Gauge
Manufacturer : Pudenberg
Model : EN837-1
Serial Number : 97317026
Type : Analog
Range : 0 to 10 bar
Calibrated Range : 0 to 10 bar
Resolution : 0.2 bar
Accuracy Class : 1.0
Tolerance : ± 1 % of Full Scale
As found : In Tolerance
Calibrated By : Asjad Rafiq
Calibration Date : November 3, 2024
Calibration Due : 02 November 2025
Last Calibration : NA



DEVIATION FROM STANDARD METHOD : None

REMARK (S) :

1. No adjustments were made prior to calibration.
2. Overload test was not performed.
3. No accessories were fitted.
4. Values were rounded in computation.

ENVIRONMENTAL CONDITIONS DURING TEST

Ambient Temperature : $25.0 \pm 2^{\circ}\text{C}$ Humidity : $44.0 \pm 5\% \text{ RH}$ Atm. Pressure : $1011.0 \pm 10 \text{ hPa}$

CALIBRATION METHOD

The above equipment has been calibrated in accordance with Calibration Procedure QC/CP/R/03 Rev. 01

The above instrument has been calibrated in increasing and decreasing pressure mode using a Digital Pressure Calibrator as the reference standard having traceability to Internationally recognised standards.

REFERENCE EQUIPMENT USED

DESCRIPTION	SERIAL NO.	MAKE / MODEL	CALIBRATION DATE	CALIBRATION DUE DATE
Digital Pressure Gauge	468802	AMETEK / XP2i	August 10, 2024	August 9, 2025

CALIBRATION TEST RESULTS

Measured Values on Unit Under Calibration (M)	Ref. Standard Values (A)	Deviation (D=M-A)	Error Percentage
bar	bar	bar	$\pm$
0	0.00	0.00	0.0
2	1.09	0.91	9.1
4	4.10	-0.10	-1.0
8	8.10	-0.10	-1.0
10	10.10	-0.10	-1.0

Calibration results were found to conform as per specified accuracy requirements. Above Instrument has **PASSED** its Calibration.

- ☒ The results are as found (no adjustment done).
☐ The results are post adjustment.