


# Calibration Certificate

Al Takamul Yard North Rumailah, Iraq

• Phone : +964 7810009138 • www.qualitycontrol-iraq.com • E-mail: op@qualitycontrol-iraq

**Date of Issue:** November 12, 2024

Page 1 of 5

REQUEST NUMBER	: BY MAIL	APPROVED BY:  ABINCAH ALI QC
JOB NUMBER	: QC-YB-240044	
CERTIFICATE NUMBER	: <b>QC-YB-240044-05</b>	

### CUSTOMER DETAILS

**Name** : Halliburton Worldwide IRAQ  
**Department** : IEM  
**Address** : Western Burjesia, Oil Street, Zubair-South Iraq

### EQUIPMENT IDENTIFICATION AND SPECIFICATIONS

**Description** : **Clamp Meter**  
**Type of Indication** : Digital  
**Manufacturer** : Fluke, USA  
**Model** : 376  
**Serial Number** : 31140396WS  
**SAP** : 300094872



#### Calibrated Range:

Voltage (DC)	0.1 mV	to	900 V
Voltage (AC) @ 50 Hz	0.1 mV	to	900 V
Voltage (AC) @ 1 kHz	0.1 mV	to	900 V
Current (DC)	0.1 A	to	10 A
Current (AC) @ 50 Hz	0.1 A	to	10 A
Current (AC) @ 1 kHz	0.1 A	to	10 A
Resistance	1 Ω	to	60 K Ω

#### Resolution:

Voltage (DC)	0.1 mV	0.1 mV	1 V
Voltage (AC) @ 50 Hz	0.1 mV	0.1 mV	1 V
Voltage (AC) @ 1 kHz	0.1 mV	0.1 mV	1 V
Current (DC)	0.1 mA	0.001 mA	0.001 A
Current (AC) @ 50 Hz	0.1 mA	0.001 mA	0.001 A
Current (AC) @ 1 kHz	0.1 mA	0.001 mA	0.001 A
Resistance	0.1 Ω	1 Ω	10 Ω

**As found** : In Tolerance  
**Calibrated By** : Abdulrahman Loay  
**Calibration Date** : **November 12, 2024**  
**Calibration Due** : **November 11, 2025**  
**Last Calibration Date** : **September 19, 2023**

### ENVIRONMENTAL CONDITIONS DURING TEST

**Ambient Temperature** : 22 °C ± 2 °C  
**Relative Humidity** : 40 %RH ± 5 %RH

### CALIBRATION METHOD

The above equipment has been calibrated in accordance with QC Calibration Procedure # QC/CP/E/01  
 The deviations of the measurements obtained from UUC with respect to reference standards are determined to obtain the error.

### TRACEABILITY

The measurements made by Quality Control Labs, realize the physical units of measurements (SI), through its state of the art calibration standards that are controlled and maintained by QC.

### REFERENCE EQUIPMENT USED :

DESCRIPTION	MAKE	MODEL #	SERIAL #	CALIBRATION DATE	CALIBRATION DUE DATE
Multifunction Calibrator	Fluke, USA	5522A	2806902	8/20/2024	8/19/2025
Ref Multimeter	Fluke, USA	8508A	276568089	8/20/2024	8/19/2025
Decade Resistance Box	Corpico	RBB6-B	18F-1093	8/20/2024	8/19/2025
Programmable Inductance substitutor	IET Labs	PLS-1492	J1-1419517	8/20/2024	8/19/2025

# CERTIFICATE OF CALIBRATION

CERTIFICATE NUMBER

**QC-YB-240044-05**

REQUEST NUMBER: BY MAIL  
JOB NUMBER: QC-YB-240044

Page 2 of 5

## CALIBRATION TEST RESULTS

### Measurement Data for DC Voltage Zero or Offset Readings of UUC

Before Adjustment	After Adjustment
$\mu\text{V}$	$\mu\text{V}$
0	0

Before Adjustment	After Adjustment
mV	mV
0	0

Before Adjustment	After Adjustment
V	V
0	0

Readings on UUC	Readings on Ref. Standard	Error	Uncertainty (95 % C.L.)
$\mu\text{V}$	$\mu\text{V}$	$\mu\text{V}$	$\pm (\mu\text{V})$
99.9	100.00	-0.10	0.1
mV	mV	mV	$\pm (\text{mV})$
0.9	1.00	-0.10	0.1
9.8	10.00	-0.20	0.1
100	101.25	-1.25	0.1
500	501.05	-1.05	0.1
900	901.23	-1.23	0.1
V	V	V	$\pm (\text{V})$
1.002	1.00	0.00	0.1
-1.002	-1.00	0.00	0.1
10.01	10.05	-0.04	0.1
100	100.00	0.00	0.1
300	300.36	-0.36	0.1
500	501.23	-1.23	1
901	901.36	-0.36	1
-902	-900.00	-2.00	1

### Measurement Data for AC Voltage @ 50 Hz Zero or Offset Readings of UUC

Before Adjustment	After Adjustment
$\mu\text{V}$	$\mu\text{V}$
0	0

Before Adjustment	After Adjustment
mV	mV
0	0

Before Adjustment	After Adjustment
V	V
0	0

Readings on UUC	Readings on Ref. Standard	Error	Uncertainty (95 % C.L.)
$\mu\text{V}$	$\mu\text{V}$	$\mu\text{V}$	$\pm (\mu\text{V})$
99.9	100.000	-0.100	0.1
mV	mV	mV	$\pm (\text{mV})$
0.9	1.000	-0.100	0.1
9.8	10.000	-0.200	0.1
100	100.005	-0.005	0.1
500	499.998	0.002	0.1
900	899.998	0.002	0.1
V	V	V	$\pm (\text{V})$
1.002	0.999980	0.002020	0.1
10.01	9.99998	0.01002	0.1
100.0	100.0130	-0.0130	0.1
300.0	300.015	-0.015	0.1
500	501.256	-1.256	1
901	900.080	0.920	1

**CERTIFICATE OF CALIBRATION**

CERTIFICATE NUMBER

**QC-YB-240044-05**

REQUEST NUMBER: BY MAIL  
JOB NUMBER: QC-YB-240044

Page 3 of 5

**Measurement Data for AC Voltage @ 1 kHz**

Zero or Offset Readings of UUC

Before Adjustment	After Adjustment
$\mu V$	$\mu V$
0	0

Before Adjustment	After Adjustment
mV	mV
0	0

Before Adjustment	After Adjustment
V	V
0	0

Readings on UUC	Readings on Ref. Standard	Error	Uncertainty (95 % C.L)
$\mu V$	$\mu V$	$\mu V$	$\pm (\mu V)$
99.9	100.00	-0.10	0.1
mV	mV	mV	$\pm (mV)$
0.9	1.001	-0.101	0.1
9.8	10.001	-0.201	0.1
100	100.001	-0.001	0.1
500	500.000	0.000	0.1
900	900.000	0.000	0.1
V	V	V	$\pm (V)$
1.002	1.000010	0.001990	0.1
10.01	10.00001	0.00999	0.1
100.0	99.9870	0.0130	0.1
300.0	299.987	0.013	0.1
500	499.987	0.013	1
901	899.930	1.070	1

**Measurement Data for DC Current**

Zero or Offset Readings of UUC

Before Adjustment	After Adjustment
$\mu A$	$\mu A$
0	0

Before Adjustment	After Adjustment
$\mu A$	$\mu A$
0	0

Before Adjustment	After Adjustment
$\mu A$	$\mu A$
0	0

Readings on UUC	Readings on Ref. Standard	Error	Uncertainty (95 % C.L)
$\mu A$	$\mu A$	$\mu A$	$\pm (\mu A)$
99.9	99.999	-0.099	0.1
mA	mA	mA	$\pm (mA)$
0.9	1.000	-0.100	0
9.8	10.000	-0.200	0.1
100	99.999	0.001	0.1
500	499.994	0.006	0.1
900	900.010	-0.010	0.1
A	A	A	$\pm (A)$
1.002	1.000010	0.001990	0.01
-1.002	-1.000030	-0.001970	0.01
3.002	2.999950	0.002050	0.01
5.003	4.999950	0.003050	0.06
9.003	8.999800	0.003200	0.06

**CERTIFICATE OF CALIBRATION**

CERTIFICATE NUMBER

**QC-YB-240044-05**

REQUEST NUMBER: BY MAIL  
JOB NUMBER: QC-YB-240044

Page 4 of 5

**Measurement Data for AC Current @ 50 Hz**

Zero or Offset Readings of UUC

Before Adjustment	After Adjustment
μA	μA
0	0

Before Adjustment	After Adjustment
μA	μA
0	0

Before Adjustment	After Adjustment
μA	μA
0	0

Readings on UUC	Readings on Ref. Standard	Error	Uncertainty (95 % C.L)
μA	μA	μA	± (μA)
99.9	99.999	-0.099	0.1
mA	mA	mA	± (mA)
0.9	1.000	-0.100	0
9.8	10.000	-0.200	0.1
100	100.000	0.000	0.1
500	500.030	-0.030	0.1
900	899.900	0.100	0.2
A	A	A	± (A)
1.002	1.000100	0.001900	0.01
3.002	3.000040	0.001960	0.01
5.003	5.000040	0.002960	0.06
9.003	9.001000	0.002000	0.06

**Measurement Data for AC Current @ 1 kHz**

Zero or Offset Readings of UUC

Before Adjustment	After Adjustment
μA	μA
0	0

Before Adjustment	After Adjustment
μA	μA
0	0

Before Adjustment	After Adjustment
μA	μA
0	0

**Measurement Data for AC Current @ 1 kHz**

Readings on UUC	Readings on Ref. Standard	Error	Uncertainty (95 % C.L)
μA	μA	μA	± (μA)
99.9	99.999	-0.099	0.1
mA	mA	mA	± (mA)
0.9	1.000	-0.100	0
9.8	10.000	-0.200	0.1
100	100.020	-0.020	0.1
500	500.040	-0.040	0.1
900	900.050	-0.050	0.2
A	A	A	± (A)
1.002	0.999950	0.002050	0.01
3.002	2.999780	0.002220	0.01
5.003	4.999780	0.003220	0.06
9.003	9.001000	0.002000	0.06

**CERTIFICATE OF CALIBRATION**

CERTIFICATE NUMBER

**QC-YB-240044-05**

REQUEST NUMBER: BY MAIL  
JOB NUMBER: QC-YB-240044

Page 5 of 5

**Measurement Data for Resistance**

Zero or Offset Readings of UUC

Before Adjustment	After Adjustment
$\Omega$	$\Omega$
0	0

Before Adjustment	After Adjustment
k $\Omega$	k $\Omega$
0	0

Before Adjustment	After Adjustment
M $\Omega$	M $\Omega$
0	0



Readings on UUC	Readings on Ref. Standard	Error	Uncertainty (95 % C.L)
$\Omega$	$\Omega$	$\Omega$	$\pm (\Omega)$
1.1	1.00010	0.09990	0.08
10.1	9.99990	0.10010	0.10
100.1	99.9984	0.1016	0.08
600.3	600.2300	0.0700	0.08
k $\Omega$	k $\Omega$	k $\Omega$	$\pm (k\Omega)$
1.003	1.000000	0.003000	0.01
10.002	9.99996	0.00204	0.01
20.0	20.0045	-0.0045	0.06
40.28	40.2324	0.0476	0.05
60.00	59.995	0.005	0.08

Calibration results were found to conform as per specified accuracy requirements. Above Instrument has **PASSED** its Calibration.

The reported expanded uncertainty is based on a standard uncertainty multiplied by a coverage factor k=2, providing a level of confidence of approximately 95%. The uncertainty evaluation has been carried out in accordance with international practice.

DEVIATION FROM STANDARD METHOD : None

REMARK (S) :  The results are as found (no adjustment done).  
 The results are post adjustment.

CALIBRATED BY	REVIEWED & APPROVED BY LAB INCHARGE	CLIENT
 Abdulrahman Loay	 LAB INCHARGE Asjad Rafiq	

*This certificate is issued in accordance with the laboratory accreditation requirements as per ISO/IEC17025:2017.*

*This certificate may not be reproduced other than in full, unless permission for the publication of an approved extract has been obtained in writing from the director of QC. It does not of itself impute to the subject of the calibration any attributes beyond those shown by the data contained herein.*

~ END OF DOCUMENT ~