

Calibration Certificate

Al Takamul Yard North Rumailah, Iraq

• Phone : +964 7810009138 • www.qualitycontrol-iraq.com • E-mail: op@qualitycontrol-iraq

Date of Issue: November 12, 2024

Page 1 of 5

| | |
|---|-----------------------------|
| REQUEST NUMBER : By Mail | APPROVED BY LAB INCHARGE QC |
| JOB NUMBER : QC-YB-240044 | |
| CERTIFICATE NUMBER : QC-YB-240044-06 | Asjad Rafiq |

CUSTOMER DETAILS

Name : **Halliburton Worldwide IRAQ**
 Department : IEM
 Address : Western Burjesia, Oil Street, Zubair, South Iraq

EQUIPMENT IDENTIFICATION AND SPECIFICATIONS

Description : **Multimeter**
 Type of Indication : Digital
 Manufacturer : Fluke, USA
 Model : 771V
 Serial Number : 23490119
 SAP No. : 300094893



Calibrated Range:

| | | | |
|----------------------|---------|----|----------|
| Voltage (DC) | 0.1 mV | to | 1000 V |
| Voltage (AC) @ 50 Hz | 1 mV | to | 1000 V |
| Voltage (AC) @ 1 kHz | 1 mV | to | 1000 V |
| Current (DC) | 0.01 mA | to | 10 A |
| Current (AC) @ 50 Hz | 0.01 mA | to | 10 A |
| Current (AC) @ 1 kHz | 0.01 mA | to | 10 A |
| Resistance | 0.1 ohm | to | 50 M-ohm |

Resolution:

| | | | | | |
|----------------------|---------|-------------|-------------|------|----|
| Voltage (DC) | 0.1 mV | 0.001 V | 0.01 V | 0.1V | 1V |
| Voltage (AC) @ 50 Hz | 0.1 mV | 0.001 V | 0.01 V | 0.1V | 1V |
| Voltage (AC) @ 1 kHz | 0.1 mV | 0.001 V | 0.01 V | 0.1V | 1V |
| Current (DC) | 0.1 mA | 0.001 mA | 0.001 A | | |
| Current (AC) @ 50 Hz | 0.1 mA | 0.001 mA | 0.001 A | | |
| Current (AC) @ 1 kHz | 0.1 mA | 0.001 mA | 0.001 A | | |
| Resistance | 0.1 ohm | 0.001 k-ohm | 0.001 M-ohm | | |

As found condition : In Tolerance

Calibration Date : November 12, 2024

Calibration Due : November 11, 2025

Last Calibration Date : September 19, 2023

ENVIRONMENTAL CONDITIONS DURING TEST

Ambient Temperature : 22 °C ± 2 °C
 Relative Humidity : 40 %RH ± 5 %RH

CALIBRATION METHOD

The above equipment has been calibrated in accordance with QC Calibration Procedure # QC/CP/E/01

The deviations of the measurements obtained from UUC with respect to reference standards are determined to obtain the error.

TRACEABILITY

The measurements made by Quality Control Labs, realize the physical units of measurements (SI), through its state of the art calibration standards that are controlled and maintained by QC.

REFERENCE EQUIPMENT USED :

| DESCRIPTION | MAKE | MODEL # | SERIAL # | CALIBRATION DATE | CALIBRATION DUE DATE |
|-------------------------------------|------------|----------|------------|------------------|----------------------|
| Multifunction Calibrator | Fluke, USA | 5522A | 2806902 | 8/20/2024 | 8/19/2025 |
| Ref Multimeter | Fluke, USA | 8508A | 276568089 | 8/20/2024 | 8/19/2025 |
| Decade Resistance Box | Corpico | RBB6-B | 18F-1093 | 8/20/2024 | 8/19/2025 |
| Programmable Inductance substitutor | IET Labs | PLS-1492 | J1-1419517 | 8/20/2024 | 8/19/2025 |

CERTIFICATE OF CALIBRATION

CERTIFICATE NUMBER

 REQUEST NUMBER: By Mail
 JOB NUMBER: QC-YB-240044

QC-YB-240044-06

Page 2 of 5

CALIBRATION TEST RESULTS
Measurement Data for DC Voltage

Zero or Offset Readings of UUC

| Before Adjustment | After Adjustment |
|-------------------|------------------|
| μV | μV |
| 0 | 0 |

| Before Adjustment | After Adjustment |
|-------------------|------------------|
| mV | mV |
| 0 | 0 |

| Before Adjustment | After Adjustment |
|-------------------|------------------|
| V | V |
| 0 | 0 |

| Readings on UUC | Readings on Ref. Standard | Error | Uncertainty (95 % C.L) |
|-----------------|---------------------------|-------|------------------------|
| mV | mV | mV | \pm (mV) |
| 0.9 | 1.01 | -0.11 | 0.1 |
| 9.9 | 10.00 | -0.10 | 0.1 |
| 101 | 100.00 | 1.00 | 0.1 |
| 501 | 500.00 | 1.00 | 0.1 |
| 599 | 600.03 | -1.03 | 0.1 |
| V | V | V | \pm (V) |
| 1.002 | 1.00 | 0.00 | 0.1 |
| -1.001 | -1.00 | 0.00 | 0.1 |
| 9.97 | 10.00 | -0.03 | 0.1 |
| 100 | 100.00 | 0.00 | 0.1 |
| 301 | 300.07 | 0.93 | 0.1 |
| 501 | 500.04 | 0.96 | 0.2 |
| 899 | 900.05 | -1.05 | 0.2 |
| 999 | 1000.08 | -1.08 | 0.2 |

Measurement Data for AC Voltage @ 50 Hz

Zero or Offset Readings of UUC

| Before Adjustment | After Adjustment |
|-------------------|------------------|
| μV | μV |
| 0 | 0 |

| Before Adjustment | After Adjustment |
|-------------------|------------------|
| mV | mV |
| 0 | 0 |

| Before Adjustment | After Adjustment |
|-------------------|------------------|
| V | V |
| 0 | 0 |

| Readings on UUC | Readings on Ref. Standard | Error | Uncertainty (95 % C.L) |
|-----------------|---------------------------|----------|------------------------|
| mV | mV | mV | \pm (mV) |
| 0.9 | 1.000 | -0.100 | 0.1 |
| 9.8 | 10.000 | -0.200 | 0.1 |
| 100 | 100.000 | 0.000 | 0.1 |
| 500 | 499.998 | 0.002 | 0.1 |
| 600 | 599.583 | 0.417 | 0.1 |
| V | V | V | \pm (V) |
| 1.002 | 0.99998 | 0.002020 | 0.1 |
| 10.01 | 9.99978 | 0.01022 | 0.1 |
| 100.0 | 100.0850 | -0.0850 | 0.1 |
| 300.0 | 300.013 | -0.013 | 0.1 |
| 500 | 500.013 | -0.013 | 0.1 |
| 901 | 900.080 | 0.920 | 0.2 |
| 1000 | 1000.087 | -0.087 | 0.2 |

CERTIFICATE OF CALIBRATION

CERTIFICATE NUMBER

QC-YB-240044-06

 REQUEST NUMBER: By Mail
 JOB NUMBER: QC-YB-240044

Page 3 of 5

Measurement Data for AC Voltage @ 1 kHz

Zero or Offset Readings of UUC

| Before Adjustment | After Adjustment |
|-------------------|------------------|
| μV | μV |
| 0 | 0 |

| Before Adjustment | After Adjustment |
|-------------------|------------------|
| mV | mV |
| 0 | 0 |

| Before Adjustment | After Adjustment |
|-------------------|------------------|
| V | V |
| 0 | 0 |

| Readings on UUC | Readings on Ref. Standard | Error | Uncertainty (95 % C.L) |
|-----------------|---------------------------|----------|------------------------|
| mV | mV | mV | \pm (mV) |
| 0.9 | 1.001 | -0.101 | 0.1 |
| 9.8 | 10.003 | -0.203 | 0.1 |
| 100 | 100.003 | -0.003 | 0.1 |
| 499 | 500.000 | -1.000 | 0.1 |
| 600 | 601.200 | -1.200 | 0.1 |
| V | V | V | \pm (V) |
| 1.001 | 1.000010 | 0.001990 | 0.1 |
| 10.01 | 10.00001 | 0.00999 | 0.1 |
| 100.0 | 99.9870 | 0.01300 | 0.1 |
| 300.0 | 299.953 | 0.04700 | 0.1 |
| 501 | 499.987 | 1.01300 | 0.2 |
| 901 | 899.930 | 1.07000 | 0.2 |
| 1000 | 1001.243 | -1.243 | 0.3 |

Measurement Data for DC Current

Zero or Offset Readings of UUC

| Before Adjustment | After Adjustment |
|-------------------|------------------|
| μA | μA |
| 0 | 0 |

| Before Adjustment | After Adjustment |
|-------------------|------------------|
| μA | μA |
| 0 | 0 |

| Before Adjustment | After Adjustment |
|-------------------|------------------|
| μA | μA |
| 0 | 0 |

| Readings on UUC | Readings on Ref. Standard | Error | Uncertainty (95 % C.L) |
|-----------------|---------------------------|-----------|------------------------|
| mA | mA | mA | \pm (mA) |
| 0.9 | 1.000 | -0.100 | 0 |
| 9.8 | 10.000 | -0.200 | 0.1 |
| 100 | 99.999 | 0.001 | 0.1 |
| 500 | 499.987 | 0.013 | 0.1 |
| 900 | 900.010 | -0.010 | 0.1 |
| A | A | A | \pm (A) |
| 1.002 | 1.000010 | 0.001990 | 0.01 |
| -1.002 | -1.000030 | -0.001970 | 0.01 |
| 3.001 | 2.999950 | 0.001050 | 0.01 |
| 5.003 | 4.996940 | 0.006060 | 0.06 |
| 9.006 | 8.955800 | 0.050200 | 0.06 |

CERTIFICATE OF CALIBRATION

CERTIFICATE NUMBER

QC-YB-240044-06

REQUEST NUMBER: By Mail
 JOB NUMBER: QC-YB-240044

Page 4 of 5

Measurement Data for AC Current @ 50 Hz

Zero or Offset Readings of UUC

| Before Adjustment | After Adjustment |
|-------------------|------------------|
| μA | μA |
| 0 | 0 |

| Before Adjustment | After Adjustment |
|-------------------|------------------|
| μA | μA |
| 0 | 0 |

| Before Adjustment | After Adjustment |
|-------------------|------------------|
| μA | μA |
| 0 | 0 |

| Readings on UUC | Readings on Ref. Standard | Error | Uncertainty (95 % C.L) |
|-----------------|---------------------------|----------|------------------------|
| mA | mA | mA | \pm (mA) |
| 0.9 | 1.000 | -0.100 | 0 |
| 9.8 | 10.000 | -0.200 | 0.1 |
| 100 | 100.000 | 0.000 | 0.1 |
| 500 | 500.020 | -0.020 | 0.1 |
| 900 | 899.930 | 0.070 | 0.2 |
| A | A | A | \pm (A) |
| 1.002 | 1.000100 | 0.001900 | 0.01 |
| 3.003 | 3.000540 | 0.002460 | 0.01 |
| 5.002 | 5.000480 | 0.001520 | 0.06 |
| 9.002 | 9.001000 | 0.001000 | 0.06 |

Measurement Data for AC Current @ 1 kHz

Zero or Offset Readings of UUC

| Before Adjustment | After Adjustment |
|-------------------|------------------|
| μA | μA |
| 0 | 0 |

| Before Adjustment | After Adjustment |
|-------------------|------------------|
| μA | μA |
| 0 | 0 |

| Before Adjustment | After Adjustment |
|-------------------|------------------|
| μA | μA |
| 0 | 0 |

Measurement Data for AC Current @ 1 kHz

| Readings on UUC | Readings on Ref. Standard | Error | Uncertainty (95 % C.L) |
|-----------------|---------------------------|----------|------------------------|
| mA | mA | mA | \pm (mA) |
| 0.9 | 1.000 | -0.100 | 0 |
| 9.8 | 10.000 | -0.200 | 0.1 |
| 100 | 100.020 | -0.020 | 0.1 |
| 500 | 500.040 | -0.040 | 0.1 |
| 900 | 900.050 | -0.050 | 0.2 |
| A | A | A | \pm (A) |
| 1.002 | 0.997950 | 0.004050 | 0.01 |
| 3.002 | 2.999770 | 0.002230 | 0.01 |
| 5.002 | 4.999750 | 0.002250 | 0.06 |
| 9.002 | 9.001020 | 0.000980 | 0.06 |

CERTIFICATE OF CALIBRATION

CERTIFICATE NUMBER

QC-YB-240044-06

REQUEST NUMBER: By Mail
 JOB NUMBER: QC-YB-240044

Page 5 of 5

Measurement Data for Resistance

Zero or Offset Readings of UUC

| Before Adjustment | After Adjustment |
|-------------------|------------------|
| Ω | Ω |
| 0 | 0 |

| Before Adjustment | After Adjustment |
|-------------------|------------------|
| k Ω | k Ω |
| 0 | 0 |

| Before Adjustment | After Adjustment |
|-------------------|------------------|
| M Ω | M Ω |
| 0 | 0 |

| Readings on UUC | Readings on Ref. Standard | Error | Uncertainty (95 % C.L) |
|-----------------|---------------------------|------------|------------------------|
| Ω | Ω | Ω | $\pm (\Omega)$ |
| 1.1 | 1.00010 | 0.09990 | 0.08 |
| 10.1 | 9.99990 | 0.10010 | 0.10 |
| 100.2 | 99.9990 | 0.2010 | 0.08 |
| 900.2 | 900.0000 | 0.2000 | 0.08 |
| k Ω | k Ω | k Ω | $\pm (k\Omega)$ |
| 1.002 | 1.000000 | 0.002000 | 0.01 |
| 10.002 | 9.99985 | 0.00215 | 0.01 |
| 100.1 | 100.0006 | 0.0994 | 0.06 |
| 500.1 | 499.9993 | 0.1007 | 0.05 |
| 900.2 | 899.999 | 0.201 | 0.08 |
| M Ω | M Ω | M Ω | $\pm (M\Omega)$ |
| 1.001 | 1.000001 | 0.000999 | 0.12 |
| 10.000 | 10.00025 | -0.00025 | 0.01 |
| 30.003 | 29.9750 | 0.0280 | 0.02 |
| 50.002 | 49.9852 | 0.0168 | 0.02 |

Calibration results were found to conform as per specified accuracy requirements. Above Instrument has **PASSED** its Calibration.

The reported expanded uncertainty is based on a standard uncertainty multiplied by a coverage factor k=2, providing a level of confidence of approximately 95%. The uncertainty evaluation has been carried out in accordance with international practice.

DEVIATION FROM STANDARD METHOD : None

REMARK (S) :

- The results are as found (no adjustment done).
- The results are post adjustment.

| | | |
|--|--|--------|
| CALIBRATED BY  Abdulrahman Loay | REVIEWED & APPROVED BY LAB INCHARGE  LAB INCHARGE Asjad Rafiq | CLIENT |
|--|--|--------|

This certificate is issued in accordance with the laboratory accreditation requirements as per ISO/IEC17025:2017.

This certificate may not be reproduced other than in full, unless permission for the publication of an approved extract has been obtained in writing from the director of QC. It does not of itself impute to the subject of the calibration any attributes beyond those shown by the data contained herein.

~ END OF DOCUMENT ~