


Certificate of Calibration

Al Takamul Yard, North Rumaila, Iraq
 • Phone : +964 7810009138 • www.qualitycontrol-iraq.com • E-mail: op@qualitycontrol-iraq.com

Date of Issue: November 12, 2024

Page 1 of 2

REQUEST NUMBER	: BY Mail	APPROVED BY LAB IN CHARGE QC
JOB NUMBER	: QC-YB-240044	
CERTIFICATE NUMBER	: QC-YB-240044-07	

CUSTOMER DETAILS

Name : Halliburton Worldwide IRAQ
Address : Western Burjesia, Oil Street, Zubair-South Iraq
Department : IEM

EQUIPMENT IDENTIFICATION AND SPECIFICATIONS

Description : Digital Thermohygrometer
Manufacturer : EXTECH
Model : NA
Serial Number : IEM-119
SAP No. : 300310738
Temperature Range : -10 °C to 60 °C
Calibration Range : -10 °C to 50 °C
Temp Readability : 0.1 °C
Temp Accuracy : ±1.0 °C
Humidity Range : 10 %RH to 99 %RH
Humidity Readability : 1.0 %RH
Humidity Accuracy : 5.0 %RH
As Found Condition : In Tolerance
Calibrated By : Abdulrahman Loay
Calibration Date : November 12, 2024
Calibration Due : November 11, 2025
Last Calibration Date : September 19, 2023



ENVIRONMENTAL CONDITIONS DURING TEST

Average Ambient Temperature : 22 °C ± 2 °C
Relative Humidity : 40 % RH ± 5 % RH

CALIBRATION METHOD

The above equipment has been calibrated in accordance to Calibration Procedure : QC/CP/T/08 Rev. 01
 The deviations of the measurements obtained from UUC with respect to reference standards are determined to obtain the error.

REFERENCE EQUIPMENT USED :

DESCRIPTION	MAKE / MODEL	SERIAL NO.	CALIBRATION DATE	DUE DATE
Temperature Humidity Test Chamber	LIB / TH100	1003256	15-Dec-23	14-Dec-24

