


Certificate of Calibration

Al Takamul Yard, North Rumailah South Iraq

• Phone : +964 7810009138 • www.qualitycontrol-iraq.com • E-mail: op@qualitycontrol-iraq.com

Date of Issue: November 17, 2024

REQUEST NUMBER : By Contract JOB NUMBER : QC-YB-240055 CERTIFICATE NUMBER : QC-YB-240055-03	APPROVED BY:  Page 1 of 1
--	---

CUSTOMER DETAILS

Name : SLB Basra
Address : North Rumaila Base south Iraq
Department : Testing

EQUIPMENT IDENTIFICATION AND SPECIFICATIONS

Description : Pressure Gauge
Manufacturer : FG
Model : NA
Serial Number : CD874/01
Location : ESD-CD-874
Type : Analog
Range : 0 to 150 psi
Calibrated Range : 0 to 150 psi
Resolution : 5 psi
Accuracy : 1.0
Tolerance : ± 1.0 % of Full Scale
As found : In Tolerance
Calibrated By : Hussein Alaa
Calibration Date : November 17, 2024
Calibration Due : 16 November 2025
Last Calibration Date : NA



DEVIATION FROM STANDARD METHOD : None

REMARK (S) :

- 1 No adjustments were made prior to calibration.
- 2 Overload test was not performed.
- 3 No accessories were fitted.
- 4 Values were rounded in computation.

ENVIRONMENTAL CONDITIONS DURING TEST

Ambient Temperature : 25.0 ± 2°C **Humidity** : 44.0 ± 5 % RH **Atm. Pressure** : 1011.0 ± 10 hPa

CALIBRATION METHOD

The above equipment has been calibrated in accordance with Calibration Procedure QC/CP/R/03 Rev. 01
 The above instrument has been calibrated in increasing and decreasing pressure mode using a Digital Pressure Calibrator as the reference standard having traceability to Internationally recognised standards.

REFERENCE EQUIPMENT USED

DESCRIPTION	SERIAL NO.	MAKE / MODEL	CALIBRATION DATE	CALIBRATION DUE DATE
Digital Pressure Gauge	468802	AMETEK / XP2i	5 Nov, 2024	4 Nov, 2025

CALIBRATION TEST RESULTS

Measured Values on Unit Under Calibration (M)	Ref. Standard Values (A)	Deviation (D=M-A)	Error Percentage
psi	psi	psi	±
0	0	0	0.0
50	49.94	0.06	0.0
75	75.16	-0.16	-0.1
120	120.23	-0.23	-0.2
150	150.44	-0.44	-0.3

Calibration results were found to conform as per specified accuracy requirements. Above Instrument has **PASSED** its Calibration.

- The results are as found (no adjustment done).
- The results are post adjustment.