

## Certificate of Calibration

Al Takamul Yard, North Rumailah South Iraq

• Phone : +964 7810009138 • www.qualitycontrol-iraq.com • E-mail: op@qualitycontrol-iraq.com

**Date of Issue:** November 17, 2024

<b>REQUEST NUMBER</b> : By Contract <b>JOB NUMBER</b> : QC-YB-240055 <b>CERTIFICATE NUMBER</b> : QC-YB-240055-07	APPROVED BY:  
--	---

### CUSTOMER DETAILS

**Name** : SLB Basra  
**Address** : North Rumaila Base south Iraq  
**Department** : Testing

### EQUIPMENT IDENTIFICATION AND SPECIFICATIONS

**Description** : Pressure Gauge  
**Manufacturer** : NA  
**Model** : NA  
**Serial Number** : CD874/05  
**Location** : ESD-CD-874  
**Type** : Analog  
**Range** : 0 to 160 psi  
**Calibrated Range** : 0 to 160 psi  
**Resolution** : 5 psi  
**Accuracy** : 1.0  
**Tolerance** : ± 1.0 % of Full Scale  
**As found** : In Tolerance  
**Calibrated By** : Hussein Alaa  
**Calibration Date** : November 17, 2024  
**Calibration Due** : 16 November 2025  
**Last Calibration Date** : NA



**DEVIATION FROM STANDARD METHOD** : None

**REMARK (S) :**

- 1 No adjustments were made prior to calibration.
- 2 Overload test was not performed.
- 3 No accessories were fitted.
- 4 Values were rounded in computation.

### ENVIRONMENTAL CONDITIONS DURING TEST

**Ambient Temperature** : 25.0 ± 2°C    **Humidity** : 44.0 ± 5 % RH    **Atm. Pressure** : 1011.0 ± 10 hPa

### CALIBRATION METHOD

The above equipment has been calibrated in accordance with Calibration Procedure QC/CP/R/03 Rev. 01  
 The above instrument has been calibrated in increasing and decreasing pressure mode using a Digital Pressure Calibrator as the reference standard having traceability to Internationally recognised standards.

### REFERENCE EQUIPMENT USED

DESCRIPTION	SERIAL NO.	MAKE / MODEL	CALIBRATION DATE	CALIBRATION DUE DATE
Digital Pressure Gauge	468802	AMETEK / XP2i	5 Nov, 2024	4 Nov, 2025

### CALIBRATION TEST RESULTS

Measured Values on Unit Under Calibration (M)	Ref. Standard Values (A)	Deviation (D=M-A)	Error Percentage
psi	psi	psi	±
0	0	0	0.0
40	40.09	-0.09	-0.1
80	80.15	-0.15	-0.1
120	120.29	-0.29	-0.2
160	160.41	-0.41	-0.3

Calibration results were found to conform as per specified accuracy requirements. Above Instrument has **PASSED** its Calibration.

- The results are as found (no adjustment done).
- The results are post adjustment.