

## Certificate of Calibration

Al Takamul Yard, North Rumailah South Iraq

• Phone : +964 7810009138 • www.qualitycontrol-iraq.com • E-mail: op@qualitycontrol-iraq.com

**Date of Issue:** December 31, 2024

**Page 1 of 1**

REQUEST NUMBER : By Contract	CUSTOMER DETAILS
JOB NUMBER : QC-YB-240159	Name <b>Halliburton Worldwide IRAQ (Baroid)</b>
CERTIFICATE NUMBER : <b>QC-YB-240159-03</b>	Address Western Burjesia, Oil Street, District Zubair-South Iraq

Description : **Pressure Gauge**  
 Manufacturer : BADOTHERM  
 Serial Number : ATC00201  
 Type : Analog  
 Range : 0 to 60 psi  
 Calibrated Range : 0 to 60 psi  
 Resolution : 1 psi  
 Accuracy : 1.0  
 Tolerance : ± 1.0 % of Full Scale  
 As found : In Tolerance  
 Calibration Date : December 31, 2024  
**Calibration Due : 30 December 2025**  
 Last Calibration : NA



**DEVIATION FROM STANDARD METHOD :** None

**REMARK (S) :**

1. No adjustments were made prior to calibration.
2. No accessories were fitted.
3. Values were rounded in computation.

### ENVIRONMENTAL CONDITIONS DURING TEST

Ambient Temperature : 27.0 ± 2°C Humidity : 15.0 ± 5 % RH Atm. Pressure 1013.0 ± 10 hPa

### CALIBRATION METHOD

The above equipment has been calibrated in accordance with Calibration Procedure QC/CP/R/03 Rev. 01  
 The above instrument has been calibrated in increasing and decreasing pressure mode using a Digital Pressure Calibrator as the reference standard having traceability to Internationally recognised standards.

### REFERENCE EQUIPMENT USED

DESCRIPTION	SERIAL NO.	MAKE / MODEL	CALIBRATION DATE	CALIBRATION DUE DATE
Digital Pressure Gauge	468802	AMETEK / XP2i	05 Nov, 2024	04 Nov, 2025

### CALIBRATION TEST RESULTS

Measured Values on Unit Under Calibration (M)	Ref. Standard Values (A)	Deviation (D=M-A)	Full Scale Error %
psi	psi	psi	%
0	0.00	0.00	0.0
15	15.20	-0.20	-0.3
30	30.30	-0.30	-0.5
45	45.30	-0.30	-0.5
60	60.50	-0.50	-0.8

Calibration results were found to conform as per specified accuracy requirements. Above Instrument has **PASSED** its Calibration.

- The results are as found (no adjustment done).  
 The results are post adjustment.

CALIBRATED BY	REVIEWED / APPROVED BY	LAB INCHARGE	CLIENT
 Abdulrahman Loay	 Asjad Rafiq	 LAB INCHARGE	