

Certificate of Calibration (Rejected)

Al Takamul Yard, North Rumailah South Iraq

• Phone : +964 7810009138 • www.qualitycontrol-iraq.com • E-mail: op@qualitycontrol-iraq.com

Date of Issue: December 31, 2024

Page 1 of 1

REQUEST NUMBER : By Contract		CUSTOMER DETAILS
JOB NUMBER : QC-YB-240159		Name Halliburton Worldwide IRAQ (Baroid)
CERTIFICATE NUMBER : QC-YB-240159-04		Address Western Burjesia, Oil Street, District Zubair-South Iraq

Description : **Pressure Gauge**
 Serial Number : ATC00201
 Type : Analog
 Range : 0 to 60 psi
 Calibrated Range : 0 to 60 psi
 Resolution : 1 psi
 Accuracy : NA
 Tolerance : ± 1.0 % of Full Scale
 As found : Out of Tolerance
 Calibration Date : December 31, 2024
Calibration Due : NA
 Last Calibration : NA



DEVIATION FROM STANDARD METHOD : None

REMARK (S) :

1. No adjustments were made prior to calibration.
2. No accessories were fitted.
3. Values were rounded in computation.

ENVIRONMENTAL CONDITIONS DURING TEST

Ambient Temperature : 27.0 ± 2°C Humidity : 15.0 ± 5 % RH Atm. Pressure : 1013.0 ± 10 hPa

CALIBRATION METHOD

The above equipment has been calibrated in accordance with Calibration Procedure QC/CP/R/03 Rev. 01
 The above instrument has been calibrated in increasing and decreasing pressure mode using a Digital Pressure Calibrator as the reference standard having traceability to Internationally recognised standards.

REFERENCE EQUIPMENT USED



DESCRIPTION	SERIAL NO.	MAKE / MODEL	CALIBRATION DATE	CALIBRATION DUE DATE
Digital Pressure Gauge	468802	AMETEK / XP2i	05 Nov, 2024	04 Nov, 2025

CALIBRATION TEST RESULTS

Measured Values on Unit Under Calibration (M)	Ref. Standard Values (A)	Deviation (D=M-A)	Full Scale Error %
psi	psi	psi	%
0	0.00	0.00	0.0
15	21.30	-6.30	-10.5
30	36.20	-6.20	-10.3
45	50.40	-5.40	-9.0
60	66.80	-6.80	-11.3

Calibration results were not found to conform as per specified accuracy requirements. Above Instrument has **FAILED** its Calibration.

- The results are as found (no adjustment done).
 The results are post adjustment.

CALIBRATED BY  Abdulrahman Loay	REVIEWED & APPROVED BY LAB INCHARGE  LAB INCHARGE Asjad Rafiq	CLIENT
---	--	---------------