

Certificate of Calibration

Al Takamul Yard, North Rumailah South Iraq


• Phone : +964 7810009138 • www.qualitycontrol-iraq.com • E-mail: op@qualitycontrol-iraq.com

Date of Issue: March 9, 2025

Page 1 of 1

REQUEST NUMBER : By Contract	CUSTOMER DETAILS
JOB NUMBER : QC-CAL- 25089	Name : Halliburton Worldwide IRAQ (Sperry)
CERTIFICATE NUMBER : QC-CAL- 25089-07	Address : Western Burjesia, Oil Street, District Zubair-South Iraq

EQUIPMENT IDENTIFICATION AND SPECIFICATIONS

Description : Pressure / Vacuum Gauge Manufacturer : WIKA Model : NA Serial Number : IC2000 SAP : 300320928 Type : Analog Pressure Range : 0 to 200 psi Resolution : 5 psi Vacuum Range : 0 to 30 in.Hg Resolution : 5 in.Hg Accuracy : NA Tolerance : ± 1.0 % from Each Reading As found : In Tolerance	Calibration Date : March 9, 2025 Calibration Due : 08 March 2026 Last Calibration : March 11, 2024		DEVIATION FROM STANDARD METHOD : None REMARK (S) : 1. No adjustments were made prior to calibration. 2. No accessories were fitted. 3. Values were rounded in computation.
---	--	---	--

ENVIRONMENTAL CONDITIONS DURING TEST

Ambient Temperature : 22.0 ± 2°C Humidity : 44.0 ± 5 % RH Atm. Pressure : 1011.0 ± 10 hPa

CALIBRATION METHOD

The above equipment has been calibrated in accordance with Calibration Procedure QC/CP/R/03 Rev. 01
 The above instrument has been calibrated in increasing and decreasing pressure mode using a Digital Pressure Calibrator as the reference standard having traceability to Internationally recognised standards.

REFERENCE EQUIPMENT USED

DESCRIPTION	SERIAL NUMBER	MAKE/MODEL	CALIBRATION DATE	DUE DATE
Digital Compound Gauge	1504119264	ASHCROFT / 6833407	05 Nov, 2024	04 Nov, 2025

CALIBRATION TEST RESULTS

Measured Values on Unit Under Calibration (M)	Ref. Standard Values (A)	Deviation (D=M-A)	Full Scale Error %
psi	psi	psi	%
0	0.00	0.00	0.0
50	50.40	-0.40	-0.2
150	151.10	-1.10	-0.5
200	201.30	-1.30	-0.7
InHg	InHg	InHg	InHg
0	0.00	0.00	0.00
-10	-10.02	0.02	0.1
-20	-20.05	0.05	0.2
-30	-30.07	0.07	0.2

Calibration results were found to conform as per specified accuracy requirements. Above Instrument has **PASSED** its Calibration.

- The results are as found (no adjustment done).
 The results are post adjustment.

CALIBRATED BY	REVIEWED & APPROVED BY / LAB INCHARGE	CLIENT
 Abdulrahman Loay	 ASJAD RAFIQ LAB INCHARGE	