

Certificate of Calibration

Al Takamul Yard, North Rumailah South Iraq

• Phone : +964 7810009138 • www.qualitycontrol-iraq.com • E-mail: op@qualitycontrol-iraq.com

Date of Issue: March 9, 2025

Page 1 of 1

REQUEST NUMBER : By Contract		CUSTOMER DETAILS
JOB NUMBER : QC-CAL- 25089	Name : Halliburton Worldwide IRAQ (Sperry)	
CERTIFICATE NUMBER : QC-CAL- 25089-08	Address : Western Burjesia, Oil Street, District Zubair-South Iraq	

EQUIPMENT IDENTIFICATION AND SPECIFICATIONS

Description : **Pressure Gauge**
 Manufacturer : DE WIT
 Model : NA
 Serial Number : **10.020005-06**
 SAP : **300163461**
 Type : Analog
 Range : 0 to 5000 psi
 Calibrated Range : 0 to 5000 psi
 Resolution : 50 psi
 Accuracy : 1.0
 Tolerance : ± 1.0 % from Each Reading
 As found : In Tolerance
 Calibration Date : March 9, 2025
Calibration Due : 08 March 2026
 Last Calibration Date : March 12, 2024



DEVIATION FROM STANDARD METHOD : None

REMARK (S) :

1. No adjustments were made prior to calibration.
2. No accessories were fitted.
3. Values were rounded in computation.

ENVIRONMENTAL CONDITIONS DURING TEST

Ambient Temperature : 22.0 ± 2°C Humidity : 30.0 ± 5 % RH Atm. Pressure : 1011.0 ± 10 hPa

CALIBRATION METHOD

The above equipment has been calibrated in accordance with Calibration Procedure QC/CP/R/03 Rev. 01

The above instrument has been calibrated in increasing and decreasing pressure mode using a Digital Pressure Calibrator as the reference standard having traceability to Internationally recognised standards.

REFERENCE EQUIPMENT USED

DESCRIPTION	SERIAL NO.	MAKE / MODEL	CALIBRATION DATE	CALIBRATION DUE DATE
Digital Pressure Gauge	468802	AMETEK / XP2i	05 Nov, 2024	04 Nov, 2025

CALIBRATION TEST RESULTS

Measured Values on Unit Under Calibration (M)	Ref. Standard Values (A)	Deviation (D=M-A)	Full Scale Error %
psi	psi	psi	%
0	0.00	0.00	0.0
1250	1255.0	-5.0	-0.1
2500	2510.0	-10.0	-0.2
3750	3762.0	-12.0	-0.2
5000	5018.0	-18.0	-0.4

Calibration results were found to conform as per specified accuracy requirements. Above Instrument has **PASSED** its Calibration.

- The results are as found (no adjustment done).
 The results are post adjustment.

CALIBRATED BY  Abdulrahman Loay	REVIEWED & APPROVED BY  ASJAD RAFIQ LAB INCHARGE	CLIENT
--	--	--------

