

Certificate of Calibration

AlTakamul Yard, North Rumaila-South Iraq

• Phone : +964 7810009138 • www.qualitycontrol-iraq.com • E-mail: op@qualitycontrol-iraq.com

Date of Issue: May 19, 2025

Page 1 of 1

REF NUMBER : By Contract	APPROVED BY CALIBRATION LAB INCHARGE QC
DATE OF ISSUE : QC-CAL-25178	
CERTIFICATE NUMBER : QC-CAL-25178-01	

CUSTOMER DETAILS

Name : **Daqing Drilling Company Iraq**
 Address : DPAB Co.Ltd Iraq Branch Block-9
 Rig : DQ043(ML)



EQUIPMENT IDENTIFICATION AND SPECIFICATIONS

Description : **H2S Portable Sensor**
 Manufacturer : HONEYWELL
 Model : BW GASALERT H2S
 Serial Number : J620-H822854
 Type : Digital
 Calibrated Range : 25 PPM H2S
 As found : Good
 Calibrated By : Hussein Alaa
 Calibration Date : May 19, 2025
Calibration Due : November 18, 2025

Alarm Details	Low	High
25 PPM H2S	10	15



ENVIRONMENTAL CONDITIONS DURING TEST

Ambient Temperature : 22 °C ± 2 °C
 Relative Humidity : 30 %RH ± 5 %RH

CALIBRATION METHOD

A reference cylinder consisting of Sulphur Dioxide (H2S) Gas is used , which are detected by gas detector by diffusion.

TRACEABILITY

The measurements made by QC Calibration & Testing Labs are traceable to NIST, vide Calgaz Certificate of Analysis
 The measurements made by QC Calibration & Testing Labs, realize the physical units of measurements (SI), through its state of the art calibration standards that are controlled and maintained by QC.

REFERENCE EQUIPMENT USED :

DESCRIPTION	SERIAL NO.	MAKE	GAS EXPIRY DATE
Gas cylinder (25 PPM H2S)	950-106554	CALGAZ	30/06/2026

CALIBRATION TEST RESULTS

Gas Name	Values of UUC	Values on Reference Standard	Error in Reading of UUC	Uncertainty of Measurement
H2S Gas (25 PPM H2S)	25.0	25	0	± 0.2

Calibration results were found to conform as per specified accuracy requirements. Above Instrument has **PASSED** its Calibration.

- The results are as found (no adjustment done).
- The results are post adjustment.