

Certificate of Calibration

Al Takamul Yard, North Rumailah South Iraq

• Phone : +964 7810009138 • www.qualitycontrol-iraq.com • E-mail: op@qualitycontrol-iraq.com

Date of Issue: July 9, 2025

Page 1 of 1

REQUEST NUMBER	:	By Mail	CUSTOMER DETAILS	
JOB NUMBER	:	QC-CAL-25257	Name	: Halliburton Worldwide IRAQ (Sperry)
CERTIFICATE NUMBER	:	QC-CAL-25257-01	Address	: Western Burjesia, Oil Street, District Zubair-South Iraq

EQUIPMENT IDENTIFICATION AND SPECIFICATIONS

Description : **Digital Vernier Caliper**
 Manufacturer : Mitutoyo
 Model : CD-8" CSX
 Serial Number : 05136119
 SAP No : 300093875
 Type : Digital
 Range : 0 to 8 inch
 Calibration Range : 0 to 8 inch
 Readability : 0.0005 inch
 As Found : In Tolerance
 Calibration Date : July 9, 2025
Calibration Due : 08 July 2026
 Last Calibration : NA



ENVIRONMENTAL CONDITIONS DURING TEST

Ambient Temperature : 22.0 ± 2°C Humidity : 40.0 ± 5 % RH

CALIBRATION METHOD

The above equipment has been calibrated in accordance with ASME B89.1.14

The reported uncertainty is the expanded uncertainty with k=2 according to the

ISO/IEC "Guide to the expansion of uncertainty in measurement" providing a level of confidence of approximately 95%.

REFERENCE EQUIPMENT USED :



DESCRIPTION	SERIAL NO.	MAKE / MODEL	CALIBRATION DATE	DUE DATE
Gauge Block Set	87300	Mitutoyo / 99MAK004B2	November 5, 2024	November 4, 2025

CALIBRATION TEST RESULTS

Ref. Standard Values (A)	Measured Values on Unit Under Calibration (M)	Deviation (D=M-A)	Expanded Uncertainty (U)
inch	inch	inch	±inch
0.0000	0.0000	0.0000	0.000
1.0000	1.0000	0.0000	0.001
3.0000	3.0000	0.0000	0.001
6.0000	6.0000	0.0000	0.001
8.0000	8.0000	0.0000	0.001

Calibration results were found to conform as per specified accuracy requirements. Above Instrument has **PASSED** its Calibration.

- ☒ The results are as found (no adjustment done).
☐ The results are post adjustment.

CALIBRATED BY	REVIEWED & APPROVED BY LAB INCHARGE	CLIENT
 Hussein Alaa	 LAB INCHARGE Asjad Rafiq	

