

Certificate of Calibration

Al Takamul Yard, North Rumailah South Iraq

• Phone : +964 7810009138 • www.qualitycontrol-iraq.com • E-mail: op@qualitycontrol-iraq.com

Date of Issue: July 9, 2025

Page 1 of 1

REQUEST NUMBER	By Mail	CUSTOMER DETAILS
JOB NUMBER	QC-CAL-25257	Name : Halliburton Worldwide IRAQ (Sperry)
CERTIFICATE NUMBER	QC-CAL-25257-06	Address : Western Burjesia, Oil Street, District Zubair-South Iraq

EQUIPMENT IDENTIFICATION AND SPECIFICATIONS

Description : **Dial Indicator Gauge**
 Manufacturer : Starrett
 Model : 25-441
 Serial Number : **DIL-11**
 SAP No : 300098404
 Type : Analog
 Range : 0 to 1 inch
 Calibration Range : 0 to 1 inch
 Readability : 0.001"
 As Found : In Tolerance
 Calibration Date : July 9, 2025
Calibration Due : July 8, 2026
 Last Calibration : NA



ENVIRONMENTAL CONDITIONS DURING TEST

Ambient Temperature : 29.0 ± 2°C Humidity : 40.0 ± 5 % RH

CALIBRATION METHOD

The above equipment has been calibrated in accordance with ASME B89.1.10M-2001

The reported uncertainty is the expanded uncertainty with k=2 according to the

ISO/IEC "Guide to the expansion of uncertainty in measurement" providing a level of confidence of approximately 95%.

REFERENCE EQUIPMENT USED :

DESCRIPTION	SERIAL NO.	MAKE / MODEL	CALIBRATION DATE	DUE DATE
Gauge Block Set	87300	Mitutoyo / 99MAK004B2	November 15, 2024	November 14, 2025

CALIBRATION TEST RESULTS

Ref. Standard Values (A)	Measured Values on Unit Under Calibration (M)	Deviation (D=M-A)	Expanded Uncertainty (U)
inch	inch	inch	±inch
0.000	0.000	0.000	0.000
0.200	0.201	0.001	0.001
0.500	0.501	0.001	0.001
0.800	0.800	0.000	0.001
1.000	0.999	-0.001	0.001

Calibration results were found to conform as per specified accuracy requirements. Above Instrument has **PASSED** its Calibration.

- The results are as found (no adjustment done).
- The results are post adjustment.

CALIBRATED BY  Hussein Alaa	REVISED & APPROVED BY LAB INCHARGE  LAB INCHARGE	CLIENT Asjad Rafiq
----------------------------------------------------------------------------------------------------------------------	-------------------------------------------------------------------------------------------------------------------------------------------	-----------------------

