

Certificate of Calibration

Al Takamul Yard, North Rumailah South Iraq

• Phone : +964 7810009138 • www.qualitycontrol-iraq.com • E-mail: op@qualitycontrol-iraq.com

Date of Issue: July 9, 2025

Page 1 of 1

REQUEST NUMBER	:	By Mail	CUSTOMER DETAILS	
JOB NUMBER	:	QC-CAL-25257	Name	: Halliburton Worldwide IRAQ (Sperry)
CERTIFICATE NUMBER	:	QC-CAL-25257-09	Address	: Western Burjesia, Oil Street, District Zubair-South Iraq

EQUIPMENT IDENTIFICATION AND SPECIFICATIONS

Description	:	Dial Indicator Gauge
Manufacturer	:	Starrett
Model	:	25-441
Serial Number	:	DIL-14
SAP No	:	300098407
Type	:	Analog
Range	:	0 to 1 inch
Calibration Range	:	0 to 1 inch
Readability	:	0.001"
As Found	:	In Tolerance
Calibration Date	:	July 9, 2025
Calibration Due	:	July 8, 2026
Last Calibration	:	NA



ENVIRONMENTAL CONDITIONS DURING TEST

Ambient Temperature : 29.0 ± 2°C Humidity : 40.0 ± 5 % RH

CALIBRATION METHOD

The above equipment has been calibrated in accordance with ASME B89.1.10M-2001

The reported uncertainty is the expanded uncertainty with k=2 according to the

ISO/IEC "Guide to the expansion of uncertainty in measurement" providing a level of confidence of approximately 95%.

REFERENCE EQUIPMENT USED :



DESCRIPTION	SERIAL NO.	MAKE / MODEL	CALIBRATION DATE	DUE DATE
Gauge Block Set	87300	Mitutoyo / 99MAK004B2	November 15, 2024	November 14, 2025

CALIBRATION TEST RESULTS

Ref. Standard Values (A)	Measured Values on Unit Under Calibration (M)	Deviation (D=M-A)	Expanded Uncertainty (U)
inch	inch	inch	±inch
0.000	0.000	0.000	0.000
0.200	0.199	-0.001	0.001
0.500	0.498	-0.002	0.001
0.800	0.799	-0.001	0.001
1.000	1.000	0.000	0.001

Calibration results were found to conform as per specified accuracy requirements. Above Instrument has **PASSED** its Calibration.

- ☒ The results are as found (no adjustment done).
☐ The results are post adjustment.

CALIBRATED BY	REVIEWED & APPROVED BY LAB INCHARGE	CLIENT
 Hussein Alaa	 LAB INCHARGE Asjad Rafiq	

