

## Certificate of Calibration

Al Takamul Yard, North Rumailah South Iraq

• Phone : +964 7810009138 • www.qualitycontrol-iraq.com • E-mail: op@qualitycontrol-iraq.com

Date of Issue: July 9, 2025

Page 1 of 1

REQUEST NUMBER	By Mail	CUSTOMER DETAILS
JOB NUMBER	QC-CAL-25257	Name : Halliburton Worldwide IRAQ (Sperry)
CERTIFICATE NUMBER	QC-CAL-25257-10	Address : Western Burjesia, Oil Street, District Zubair-South Iraq

### EQUIPMENT IDENTIFICATION AND SPECIFICATIONS

Description : **Dial Indicator Gauge**  
 Manufacturer : Starrett  
 Model : 25-441  
 Serial Number : **DIL-28**  
 SAP No : 300179908  
 Type : Analog  
 Range : 0 to 1 inch  
 Calibration Range : 0 to 1 inch  
 Readability : 0.001"  
 As Found : In Tolerance  
 Calibration Date : July 9, 2025  
**Calibration Due : July 8, 2026**  
 Last Calibration : NA



### ENVIRONMENTAL CONDITIONS DURING TEST

Ambient Temperature : 29.0 ± 2°C Humidity : 40.0 ± 5 % RH

### CALIBRATION METHOD

The above equipment has been calibrated in accordance with ASME B89.1.10M-2001

The reported uncertainty is the expanded uncertainty with k=2 according to the

ISO/IEC "Guide to the expansion of uncertainty in measurement" providing a level of confidence of approximately 95%.

### REFERENCE EQUIPMENT USED :

DESCRIPTION	SERIAL NO.	MAKE / MODEL	CALIBRATION DATE	DUE DATE
Gauge Block Set	87300	Mitutoyo / 99MAK004B2	November 15, 2024	November 14, 2025


### CALIBRATION TEST RESULTS

Ref. Standard Values (A)	Measured Values on Unit Under Calibration (M)	Deviation (D=M-A)	Expanded Uncertainty (U)
inch	inch	inch	±inch
0.000	0.000	0.000	0.000
0.200	0.199	-0.001	0.001
0.500	0.498	-0.002	0.001
0.800	0.800	0.000	0.001
1.000	1.000	0.000	0.001

Calibration results were found to conform as per specified accuracy requirements. Above Instrument has **PASSED** its Calibration.

The results are as found (no adjustment done).

The results are post adjustment.

CALIBRATED BY   Hussein Alaa	REVIEWED & APPROVED BY LAB INCHARGE   LAB INCHARGE Asjad Rafiq	CLIENT
--	--	--------

