


Certificate of Calibration

Al Takamul Yard, North Rumailah South Iraq

• Phone : +964 7810009138 • www.qualitycontrol-iraq.com • E-mail: op@qualitycontrol-iraq.com

Date of Issue: June 12, 2025

Page 1 of 1

REQUEST NUMBER	: By Contract	APPROVED BY LAB INCHARGE OC
JOB NUMBER	: QC-CAL-25229	Asjad Al Takamul Company
CERTIFICATE NUMBER	: QC-CAL-25229-46	

CUSTOMER DETAILS

Name : **Daqing Drilling Company Iraq**
 Address : Basra, South Iraq
 Department : DQ Base (CMT)

EQUIPMENT IDENTIFICATION AND SPECIFICATIONS

Description : Temperature Gauge
 Manufacturer : NA
 Model : NA
 Serial Number : PG6122546
 Location : Cement Unit #5
 Type : Analog
 Range : 65 to 150 °C
 Calibrated Range : 65 to 150 °C
 Resolution : 5 °C
 Accuracy Class : 1.0
 Tolerance : 1.0 % of Full Scale
 As found : In Tolerance
 Calibrated By : Hussein Alaa
 Calibration Date : June 12, 2025
Calibration Due : June 11, 2027



DEVIATION FROM STANDARD METHOD : None

REMARK (S) :

1. No adjustments were made prior to calibration.
2. Overload test was not performed.
3. No accessories were fitted.
4. Values were rounded in computation.

ENVIRONMENTAL CONDITIONS DURING TEST

Ambient Temperature : 28.0 ± 2°C Humidity : 44.0 ± 5 % RH Atm. Pressure : 1011.0 ± 10 hPa

CALIBRATION METHOD

The above equipment has been calibrated in accordance with Calibration Procedure QC/CP/R/03 Rev. 01
 The above instrument has been calibrated in increasing and decreasing pressure mode using a Digital Pressure Calibrator as the reference standard having traceability to Internationally recognised standards.

REFERENCE EQUIPMENT USED

DESCRIPTION	SERIAL NO.	MAKE / MODEL	CALIBRATION DATE	CALIBRATION DUE DATE
Metrology well	B85655	FLUKE / 9170	August 5, 2024	August 4, 2025

CALIBRATION TEST RESULTS

Measured Values on Unit Under Calibration (M)	Ref. Standard Values (A)	Deviation (D=M-A)	Error Percentage
°C	°C	°C	%
0	0	0.00	0.0
65.0	65.05	-0.05	-0.03
90.0	90.06	-0.06	-0.04
120.0	120.05	-0.05	-0.03
150.0	150.08	-0.08	-0.05

Calibration results were found to conform as per specified accuracy requirements. Above Instrument has **PASSED** its Calibration.

- The results are as found (no adjustment done).
 The results are post adjustment.