

## Certificate of Calibration

Al Takamul Yard, North Rumailah South Iraq

• Phone : +964 7810009138 • www.qualitycontrol-iraq.com • E-mail: op@qualitycontrol-iraq.com

Date of Issue: June 19, 2025

Page 1 of 1

REQUEST NUMBER : by mail	CUSTOMER DETAILS
JOB NUMBER : QC-CAL-25240	Name <b>Halliburton Worldwide IRAQ (TSS)</b>
CERTIFICATE NUMBER : <b>QC-CAL-25240-11</b>	Address Western Burjesia, Oil Street, District Zubair-South Iraq

Description : **Pressure Gauge**  
 Manufacturer : STEWART-BUCHANAN  
 Model : NA  
 Serial Number : **ESD-PG-003**  
 Type : Analog  
 Location : ESD PANEL (SAP # 13311810)  
 Range : 0 to 6000 psi  
 Calibrated Range : 0 to 6000 psi  
 Resolution : 200 psi  
 Tolerance : ± 1.0 % of Full Scale  
 As found : In Tolerance  
 Calibration Date : June 19, 2025  
**Calibration Due : June 18, 2026**  
 Last Calibration Date : May 12, 2024



**DEVIATION FROM STANDARD METHOD : None**

**REMARK (S) :**

1. No adjustments were made prior to calibration.
2. No accessories were fitted.
3. Values were rounded in computation.

### ENVIRONMENTAL CONDITIONS DURING TEST

Ambient Temperature : 22.0 ± 2°C Humidity : 44.0 ± 5 % RH Atm. Pressure 1011.0 ± 10 hPa

### CALIBRATION METHOD

The above equipment has been calibrated in accordance with Calibration Procedure QC/CP/R/03 Rev. 01

The above instrument has been calibrated in increasing and decreasing pressure mode using a Digital Pressure Calibrator as the reference standard having traceability to Internationally recognised standards.

### REFERENCE EQUIPMENT USED

DESCRIPTION	SERIAL NO.	MAKE / MODEL	CALIBRATION DATE	CALIBRATION DUE DATE
Digital Pressure Gauge	468802	AMETEK / XP2i	5 Nov, 2024	4 Nov, 2025

### CALIBRATION TEST RESULTS

Measured Values on Unit Under Calibration (M)	Ref. Standard Values (A)	Deviation (D=M-A)	Full Scale Error %
0	0	0	%
0	0.00	0.00	0.0
1400	1410.00	-10.00	-0.2
3000	3014.00	-14.00	-0.2
4600	4618.00	-18.00	-0.3
6000	6024.00	-24.00	-0.4

Calibration results were found to conform as per specified accuracy requirements. Above Instrument has **PASSED** its Calibration.

- The results are as found (no adjustment done).  
 The results are post adjustment.

CALIBRATED BY  Abdulrahman Loay	REVISED & APPROVED BY INCHARGE  LAB INCHARGE Asjad Rafiq	CLIENT
--	--	--------