

Certificate of Calibration

Al Takamul Yard, North Rumailah South Iraq

• Phone : +964 7810009138 • www.qualitycontrol-iraq.com • E-mail: op@qualitycontrol-iraq.com

Date of Issue: July 10, 2025

Page 1 of 1

REQUEST NUMBER	: By Mail	CUSTOMER DETAILS	
JOB NUMBER	: QC-CAL-25258	Name	Halliburton Worldwide IRAQ (TSS)
CERTIFICATE NUMBER	: QC-CAL-25258-01	Address	Western Burjesia, Oil Street, District Zubair-South Iraq

Description : **N2 Pump**
 Manufacturer : MAXIMATOR
 Model : DLE75-1-UU-M
 Serial Number : **12067761**
 Type : Analog

Range : 0 to 10875 psi
 Calibrated Range : 0 to 10875 psi

Compression Ratio : 1:2 Max Drive Pressure : 145 psi

Tolerance ± 1.0 % of Full Scale
 As found : In Tolerance
 Calibration Date : July 10, 2025
Calibration Due : July 9, 2026



DEVIATION FROM STANDARD METHOD : None

REMARK (S) :

1. No adjustments were made prior to calibration.
2. No accessories were fitted.
3. Values were rounded in computation.

ENVIRONMENTAL CONDITIONS DURING TEST

Ambient Temperature : 29.0 ± 2°C Humidity : 44.0 ± 5 % RH Atm. Pressure : 1011.0 ± 10 hPa

CALIBRATION METHOD

The above equipment has been calibrated in accordance with Calibration Procedure QC/CP/R/03 Rev. 01
 The above instrument has been calibrated in increasing and decreasing pressure mode using a Digital Pressure Calibrator as the reference standard having traceability to Internationally recognised standards.

REFERENCE EQUIPMENT USED


DESCRIPTION	SERIAL NO.	MAKE / MODEL	CALIBRATION DATE	CALIBRATION DUE DATE
Digital Pressure Gauge	468802	AMETEK / XP2i	5 Nov, 2024	4 Nov, 2025

CALIBRATION TEST RESULTS

Measured Values on Unit Under Calibration (M)	Ref. Standard Values (A)	Deviation (D=M-A)	Full Scale Error %
psi	psi	psi	%
0	0.00	0.00	0.0
2750	2759.00	-9.00	-0.1
5250	5263.00	-13.00	-0.1
8250	8268.00	-18.00	-0.2
10875	10900.00	-25.00	-0.2

Calibration results were found to conform as per specified accuracy requirements. Above Instrument has **PASSED** its Calibration.

- The results are as found (no adjustment done).
 The results are post adjustment.

CALIBRATED BY	REVIEWED & LAB INCHARGE	CLIENT
 Hussein Alaa	 LAB INCHARGE Asjad Rafiq	