

Certificate of Calibration

Al Takamul Yard, North Rumailah South Iraq

• Phone : +964 7810009138 • www.qualitycontrol-iraq.com • E-mail: op@qualitycontrol-iraq.com

Date of Issue: July 10, 2025

Page 1 of 1

| | | |
|---|--|-------------------------|
| REQUEST NUMBER : By Mail | | CUSTOMER DETAILS |
| JOB NUMBER : QC-CAL-25258 | Name : Halliburton Worldwide IRAQ (TSS) | |
| CERTIFICATE NUMBER : QC-CAL-25258-09 | Address : Western Burjesia, Oil Street, District Zubair-South Iraq | |

EQUIPMENT IDENTIFICATION AND SPECIFICATIONS

Description : **Pressure Chart Recorder**
 Manufacturer : DEWIT
 Model : PP-12-T
 Serial Number : 15.00300.1-01
 Type : Analog
 Range : 0 to 15000 psi
 Calibrated Range : 0 to 15000 psi
 Resolution : 500 psi
 Accuracy : NA
 Tolerance : ± 1.0 % of Full Scale
 As found : In Tolerance
 Calibration Date : July 10, 2025
Calibration Due : July 9, 2026
 Last Calibration Date : NA



DEVIATION FROM STANDARD METHOD : None

REMARK (S) :

1. No adjustments were made prior to calibration.
2. No accessories were fitted.
3. Values were rounded in computation.

ENVIRONMENTAL CONDITIONS DURING TEST

Ambient Temperature : 22.0 ± 2°C Humidity : 30.0 ± 5 % RH Atm. Pressure : 1012.0 ± 10 hPa

CALIBRATION METHOD

The above equipment has been calibrated in accordance with Calibration Procedure QC/CP/R/05 Rev. 01

The above instrument has been calibrated in increasing and decreasing pressure mode using a Digital Pressure Calibrator as the reference standard having traceability to Internationally recognised standards.

REFERENCE EQUIPMENT USED


| DESCRIPTION | SERIAL NO. | MAKE / MODEL | CALIBRATION DATE | CALIBRATION DUE DATE |
|------------------------|------------|---------------|------------------|----------------------|
| Digital Pressure Gauge | 468802 | AMETEK / XP2i | 05 Nov, 2024 | 04 Nov, 2025 |

CALIBRATION TEST RESULTS

| Measured Values on Unit Under Calibration (M) | Ref. Standard Values (A) | Deviation (D=M-A) | Error of Full Scale % |
|---|--------------------------|-------------------|-----------------------|
| psi | psi | psi | % |
| 0 | 0.0 | 0.0 | 0.0 |
| 3750 | 3762.0 | -12.0 | -0.1 |
| 7500 | 7519.0 | -19.0 | -0.1 |
| 11000 | 11029.0 | -29.0 | -0.2 |
| 15000 | 15044.0 | -44.0 | -0.3 |

Calibration results were found to conform as per specified accuracy requirements. Above Instrument has **PASSED** its Calibration.

- The results are as found (no adjustment done).
 The results are post adjustment.

| | | |
|---|--|--------------------------|
| CALIBRATED BY  Hussein Alaa | REVIEWED & APPROVED BY LAB INCHARGE  LAB INCHARGE: Asjad Rafiq | CLIENT (Blank) |
|---|--|--------------------------|