

Certificate of Calibration

Al Takamul Yard, North Rumailah South Iraq

• Phone : +964 7810009138 • www.qualitycontrol-iraq.com • E-mail: op@qualitycontrol-iraq.com

Date of Issue: July 10, 2025

Page 1 of 1

REQUEST NUMBER : By Mail		CUSTOMER DETAILS
JOB NUMBER : QC-CAL-25258	Name : Halliburton Worldwide IRAQ (TSS)	
CERTIFICATE NUMBER : QC-CAL-25258-10	Address : Western Burjesia, Oil Street, District Zubair-South Iraq	

EQUIPMENT IDENTIFICATION AND SPECIFICATIONS

Description : **Pressure Chart Recorder**
 Manufacturer : DEWIT
 Model : PP-12-T
 Serial Number : 11.00850.1-08
 Type : Analog
 Range : 0 to 15000 psi
 Calibrated Range : 0 to 15000 psi
 Resolution : 500 psi
 Accuracy : NA
 Tolerance : ± 1.0 % of Full Scale
 As found : In Tolerance
 Calibration Date : July 10, 2025
Calibration Due : July 9, 2026
 Last Calibration Date : NA



DEVIATION FROM STANDARD METHOD : None

REMARK (S) :

1. No adjustments were made prior to calibration.
2. No accessories were fitted.
3. Values were rounded in computation.

ENVIRONMENTAL CONDITIONS DURING TEST

Ambient Temperature : 22.0 ± 2°C Humidity : 30.0 ± 5 % RH Atm. Pressure : 1012.0 ± 10 hPa

CALIBRATION METHOD

The above equipment has been calibrated in accordance with Calibration Procedure QC/CP/R/05 Rev. 01

The above instrument has been calibrated in increasing and decreasing pressure mode using a Digital Pressure Calibrator as the reference standard having traceability to Internationally recognised standards.

REFERENCE EQUIPMENT USED


DESCRIPTION	SERIAL NO.	MAKE / MODEL	CALIBRATION DATE	CALIBRATION DUE DATE
Digital Pressure Gauge	468802	AMETEK / XP2i	05 Nov, 2024	04 Nov, 2025

CALIBRATION TEST RESULTS

Measured Values on Unit Under Calibration (M)	Ref. Standard Values (A)	Deviation (D=M-A)	Error of Full Scale %
psi	psi	psi	%
0	0.0	0.0	0.0
3750	3768.0	-18.0	-0.1
7500	7524.0	-24.0	-0.2
11000	11035.0	-35.0	-0.2
15000	15051.0	-51.0	-0.3

Calibration results were found to conform as per specified accuracy requirements. Above Instrument has **PASSED** its Calibration.

- The results are as found (no adjustment done).
 The results are post adjustment.

CALIBRATED BY  Hussein Alaa	REVIEWED & APPROVED BY LAB INCHARGE  LAB INCHARGE: Asjad Rafiq	CLIENT (Blank)
---	--	--------------------------