

## Certificate of Calibration

Al Takamul Yard, North Rumailah South Iraq

• Phone : +964 7810009138 • www.qualitycontrol-iraq.com • E-mail: op@qualitycontrol-iraq.com

**Date of Issue:** July 10, 2025

Page 1 of 1

REQUEST NUMBER	: By Mail	CUSTOMER DETAILS	
JOB NUMBER	: QC-CAL-25258	Name	Halliburton Worldwide IRAQ (TSS)
CERTIFICATE NUMBER	: <b>QC-CAL-25258-14</b>	Address	Western Burjesia, Oil Street, District Zubair-South Iraq

Description : **N2 Pump**  
 Manufacturer : MAXIMATOR  
 Model : DLE75-1-UU-M  
 Serial Number : **102008826**  
 Type : Analog

---

Range : 0 to 10875 psi  
 Calibrated Range : 0 to 10875 psi

---

Compression Ratio : 1:2      Max Drive Pressure : 145 psi

Tolerance ± 1.0 % of Full Scale  
 As found : In Tolerance  
 Calibration Date : July 10, 2025  
**Calibration Due : July 9, 2026**



**DEVIATION FROM STANDARD METHOD :** None

**REMARK (S) :**

1. No adjustments were made prior to calibration.
2. No accessories were fitted.
3. Values were rounded in computation.

### ENVIRONMENTAL CONDITIONS DURING TEST

Ambient Temperature : 29.0 ± 2°C      Humidity : 44.0 ± 5 % RH      Atm. Pressure : 1011.0 ± 10 hPa

### CALIBRATION METHOD

The above equipment has been calibrated in accordance with Calibration Procedure QC/CP/R/03 Rev. 01

The above instrument has been calibrated in increasing and decreasing pressure mode using a Digital Pressure Calibrator as the reference standard having traceability to Internationally recognised standards.

### REFERENCE EQUIPMENT USED

DESCRIPTION	SERIAL NO.	MAKE / MODEL	CALIBRATION DATE	CALIBRATION DUE DATE
Digital Pressure Gauge	468802	AMETEK / XP2i	5 Nov, 2024	4 Nov, 2025


### CALIBRATION TEST RESULTS

Measured Values on Unit Under Calibration (M)	Ref. Standard Values (A)	Deviation (D=M-A)	Full Scale Error %
psi	psi	psi	%
0	0.00	0.00	0.0
2750	2760.00	-10.00	-0.1
5250	5266.00	-16.00	-0.1
8250	8272.00	-22.00	-0.2
10875	10908.00	-33.00	-0.3

Calibration results were found to conform as per specified accuracy requirements. Above Instrument has **PASSED** its Calibration.

The results are as found (no adjustment done).

The results are post adjustment.

CALIBRATED BY	REVIEWED & LAB INCHARGE	CLIENT
 Hussein Alaa	 LAB INCHARGE      Asjad Rafiq	