

## Certificate of Calibration

AlTakamul Yard, North Rumaila-South Iraq

• Phone : +964 7810009138 • www.qualitycontrol-iraq.com • E-mail: op@qualitycontrol-iraq.com


**Date of Issue:** July 13, 2025

Page 1 of 1

|                    |                   |                             |
|--------------------|-------------------|-----------------------------|
| REQUEST NUMBER     | : By Contract     | APPROVED BY LAB INCHARGE-QC |
| JOB NUMBER         | : QC-CAL-25262    |                             |
| CERTIFICATE NUMBER | : QC-CAL-25262-20 |                             |

### CUSTOMER DETAILS

Name : **Daqing Drilling Company Iraq**  
 Address : South Rumaila South Iraq  
 Rig : DQ038



### EQUIPMENT IDENTIFICATION AND SPECIFICATIONS

Description : **H2S Portable Sensor**  
 Manufacturer : HONEYWELL  
 Model : BW CLIP H2S  
 Serial Number : 5220BWC01230213134  
 Type : Digital  
 Calibrated Range : 25 PPM H2S  
 As found : Good  
 Calibrated By : Hussein Alaa  
 Calibration Date : July 13, 2025  
**Calibration Due** : **October 12, 2025**



| Alarm Details | Low | High |
|---------------|-----|------|
| H2S PPM       | 10  | 15   |

### ENVIRONMENTAL CONDITIONS DURING TEST

Ambient Temperature : 22 °C ± 2 °C  
 Relative Humidity : 30 %RH ± 5 %RH

### CALIBRATION METHOD

A reference cylinder consisting of Sulphur Dioxide (H2S) Gas is used , which are detected by gas detector by diffusion.

### TRACEABILITY

The measurements made by QC Calibration & Testing Labs are traceable to NIST, vide Calgaz Certificate of Analysis  
 The measurements made by QC Calibration & Testing Labs, realize the physical units of measurements (SI), through its state of the art calibration standards that are controlled and maintained by QC.

### REFERENCE EQUIPMENT USED :

| DESCRIPTION                                | SERIAL NO. | MAKE   | GAS EXPIRY DATE |
|--------------------------------------------|------------|--------|-----------------|
| Multigas Calibration cylinder (25 PPM H2S) | WO463909-3 | CALGAZ | July 31, 2026   |

### CALIBRATION TEST RESULTS

| Gas Name         | Values of UUC | Values on Reference Standard | Error in Reading of UUC | Uncertainty of Measurement |
|------------------|---------------|------------------------------|-------------------------|----------------------------|
| H2S Gas (25 ppm) | 25.0          | 25                           | 0                       | ± 0.2                      |

Calibration results were found to conform as per specified accuracy requirements. Above Instrument has **PASSED** its Calibration.

- ☒ The results are as found (no adjustment done).  
☐ The results are post adjustment.