

Certificate of Calibration

Al Takamul Yard, North Rumailah South Iraq

• Phone : +964 7810009138 • www.qualitycontrol-iraq.com • E-mail: op@qualitycontrol-iraq.com

Date of Issue: July 13, 2025

Page 1 of 1

REQUEST NUMBER	: By Contract	APPROVED BY LAB. No. 14/2025
JOB NUMBER	: QC-CAL-25262	Asjad Alaa
CERTIFICATE NUMBER	: QC-CAL-25262-38	QUALITY CONTROL TEST 2 For Engineering Services

CUSTOMER DETAILS

Name : Daqing Drilling Company Iraq
Address : Basra, South Iraq
Department : DQ038

EQUIPMENT IDENTIFICATION AND SPECIFICATIONS

Description : Pressure Gauge
Manufacturer : MARSH
Model : NA
Serial Number : CAL-DQ38-01
Location : Fire Extinguisher
Type : Analog
Range : 0 to 60 psi
Calibrated Range : 0 to 60 psi
Resolution : 1 psi
Accuracy Class : 1.0
Tolerance : 1.0 % of Full Scale
As found : In Tolerance
Calibrated By : Hussein Alaa
Calibration Date : July 13, 2025
Calibration Due : July 12, 2026



DEVIATION FROM STANDARD METHOD : None

REMARK (S) :

1. No adjustments were made prior to calibration.
2. Overload test was not performed.
3. No accessories were fitted.
4. Values were rounded in computation.

ENVIRONMENTAL CONDITIONS DURING TEST

Ambient Temperature : 28.0 ± 2°C **Humidity** : 44.0 ± 5 % RH **Atm. Pressure** : 1011.0 ± 10 hPa

CALIBRATION METHOD

The above equipment has been calibrated in accordance with Calibration Procedure QC/CP/R/03 Rev. 01
 The above instrument has been calibrated in increasing and decreasing pressure mode using a Digital Pressure Calibrator as the reference standard having traceability to Internationally recognised standards.

REFERENCE EQUIPMENT USED

DESCRIPTION	SERIAL NO.	MAKE / MODEL	CALIBRATION DATE	CALIBRATION DUE DATE
Digital Pressure Gauge	468802	AMETEK / XP2i	5 Nov, 2024	4 Nov, 2025

CALIBRATION TEST RESULTS

Measured Values on Unit Under Calibration (M)	Ref. Standard Values (A)	Deviation (D=M-A)	Error Percentage
psi	psi	psi	%
0	0	0.00	0.0
15.0	14.8	0.20	0.33
30.0	29.9	0.14	0.23
45.0	44.9	0.08	0.13
60.0	60.0	0.03	0.05

Calibration results were found to conform as per specified accuracy requirements. Above Instrument has **PASSED** its Calibration.

- The results are as found (no adjustment done).
 The results are post adjustment.