

## Certificate of Calibration

Al Takamul Yard, North Rumailah South Iraq

• Phone : +964 7810009138 • www.qualitycontrol-iraq.com • E-mail: op@qualitycontrol-iraq.com

**Date of Issue:** July 13, 2025

Page 1 of 1

REQUEST NUMBER	: By Contract	APPROVED BY LAB. No. 1001/2025
JOB NUMBER	: QC-CAL-25262	Asjad Alaa
CERTIFICATE NUMBER	: QC-CAL-25262-41	<b>QUALITY CONTROL</b> TEST 2 For Engineering Services

### CUSTOMER DETAILS

**Name** : Daqing Drilling Company Iraq  
**Address** : Basra, South Iraq  
**Department** : DQ038

### EQUIPMENT IDENTIFICATION AND SPECIFICATIONS

**Description** : Pressure Gauge  
**Manufacturer** : MC  
**Model** : NA  
**Serial Number** : M1110230  
**Location** : Power Tong  
**Type** : Analog  
**Range** : 0 to 3600 psi  
**Calibrated Range** : 0 to 3600 psi  
**Resolution** : 100 psi  
**Accuracy Class** : 2.5  
**Tolerance** : 2.5 % of Full Scale  
**As found** : In Tolerance  
**Calibrated By** : Hussein Alaa  
**Calibration Date** : July 13, 2025  
**Calibration Due** : July 12, 2026



**DEVIATION FROM STANDARD METHOD** : None

**REMARK (S) :**

1. No adjustments were made prior to calibration.
2. Overload test was not performed.
3. No accessories were fitted.
4. Values were rounded in computation.

### ENVIRONMENTAL CONDITIONS DURING TEST

**Ambient Temperature** : 28.0 ± 2°C    **Humidity** : 44.0 ± 5 % RH    **Atm. Pressure** : 1011.0 ± 10 hPa

### CALIBRATION METHOD

The above equipment has been calibrated in accordance with Calibration Procedure QC/CP/R/03 Rev. 01  
 The above instrument has been calibrated in increasing and decreasing pressure mode using a Digital Pressure Calibrator as the reference standard having traceability to Internationally recognised standards.

### REFERENCE EQUIPMENT USED

DESCRIPTION	SERIAL NO.	MAKE / MODEL	CALIBRATION DATE	CALIBRATION DUE DATE
Digital Pressure Gauge	468802	AMETEK / XP2i	5 Nov, 2024	4 Nov, 2025

### CALIBRATION TEST RESULTS

Measured Values on Unit Under Calibration (M)	Ref. Standard Values (A)	Deviation (D=M-A)	Error Percentage
psi	psi	psi	%
0	0	0.00	0.0
900.0	892.0	8.00	0.22
1800.0	1791.0	9.00	0.25
2700.0	2694.0	6.00	0.17
3600.0	3596.0	4.00	0.11

Calibration results were found to conform as per specified accuracy requirements. Above Instrument has **PASSED** its Calibration.

- The results are as found (no adjustment done).  
 The results are post adjustment.