


Certificate of Calibration

Al Takamul Yard, North Rumailah South Iraq

• Phone : +964 7810009138 • www.qualitycontrol-iraq.com • E-mail: op@qualitycontrol-iraq.com

Date of Issue: July 13, 2025

Page 1 of 1

REQUEST NUMBER : By Contract	APPROVED BY: 
JOB NUMBER : QC-CAL-25262	
CERTIFICATE NUMBE : QC-CAL-25262-42	

CUSTOMER DETAILS

Name : **Daqing Drilling Company Iraq**
 Address : Basra, South Iraq
 Department : DQ038

EQUIPMENT IDENTIFICATION AND SPECIFICATIONS

Description : Pressure Tong Line Pull Gauge
 Manufacturer : MC
 Model : NA
 Serial Number : 7580
 Location : Manual Tong
 Type : Analog
 Range : 0 to 20000 Pound
 Calibrated Range : 0 to 20000 Pound
 Resolution : 100 Pound
 Accuracy Class : 1.0
 Tolerance : 1.0 % of Full Scale
 As found : In Tolerance
 Calibrated By : Abdulrahman Loay
 Calibration Date : July 13, 2025
Calibration Due : July 12, 2026
 Last Calibration : NA



DEVIATION FROM STANDARD METHOD : None

REMARK (S) :

1. No adjustments were made prior to calibration.
2. Overload test was not performed.
3. No accessories were fitted.
4. Values were rounded in computation.

ENVIRONMENTAL CONDITIONS DURING TEST

Ambient Temperature 25.0 ± 2°C Humidity : 44.0 ± 5 % RH Atm. Pressure 1011.0 ± 10 hPa

CALIBRATION METHOD

The above equipment has been calibrated in accordance with Calibration Procedure QC/CP/R/03 Rev. 01

The above instrument has been calibrated in increasing and decreasing pressure mode using a Digital Pressure Calibrator as the reference standard having traceability to Internationally recognised standards.

REFERENCE EQUIPMENT USED

DESCRIPTION	SERIAL NO.	MAKE / MODEL	CALIBRATION DATE	CALIBRATION DUE DATE
Digital Pressure Gauge	468802	AMETEK / XP2i	5 Nov, 2024	4 Nov, 2025

CALIBRATION TEST RESULTS

Measured Values On UUC (M)	Ref. Standard Values (A)	Deviation (D=M-A)	Error Percentage
psi	psi	psi	±%
0	0	0.00	0.0
1250.0 (5000 Pound)	1256.0	-6.00	-0.12
2500.0 (10000 Pound)	2509.0	-9.00	-0.18
3750.0 (15000 Pound)	3764.0	-14.00	-0.28
5000.0 (20000 Pound)	5021.0	-21.00	-0.42

Calibration results were found to conform as per specified accuracy requirements. Above Instrument has **PASSED** its Calibration.

- The results are as found (no adjustment done).
 The results are post adjustment.