

# Al Takamul Company For Engineering Services Quality Control - Iraq

# Certificate of Calibration

## AlTakamul Yard, North Rumaila-South Iraq

• Phone: +964 7810009138 • www.qualitycontrol-iraq.com • E-mail: op@qualitycontrol-iraq.com

Date of Issue: July 4, 2025 Page 1 of 1 By Contract REQUEST NUMBER APPROVED BY LAB INCHARGE-QC

QC-CAL-25247 JOB NUMBER QC-CAL-25247-02 CERTIFICATE NUMBER

#### **CUSTOMER DETAILS**

**Daqing Drilling Company Iraq** Name Address South Rumaila South Iraq DQ012 (Mud Logger) Rig **EQUIPMENT IDENTIFICATION AND SPECIFICATIONS** 

**H2S Portable Sensor** Description Manufacturer HONEYWELL **BW CLIP H2S** Model

5220BWC01230213013 Serial Number

Type Digital 25 PPM H2S Calibrated Range

As found Good

Calibrated By Abdulrahman Loay

July 4, 2025 Calibration Date **Calibration Due October 3, 2025** 

Alarm Details	Low	High
H2S PPM	10	15

#### ENVIRONMENTAL CONDITIONS DURING TEST

28 °C 2°C Ambient Temperature  $\pm$ Relative Humidity 30 %RH 5 %RH  $\pm$ 

## **CALIBRATION METHOD**

A reference cylinder consisting of Sulpher Dioxide (H2S) Gas is used, which are detected by gas detector by diffusion.

#### TRACEABILITY

The measurements made by QC Calibration & Testing Labs are traceable to NIST, vide Calgaz Certificate of Analysis The measurements made by QC Calibration & Testing Labs, realize the physical units of measurements (SI), through its state of the art calibration standards that are controlled and maintained by QC.

## REFERENCE EQUIPMENT USED:

DESCRIPTION	SERIAL NO.	MAKE	GAS EXPIRY DATE	
Gas Calibration cylinder (25 PPM H2S)	C011126	CALGAZ	April 30, 2026	

# **CALIBRATION TEST RESULTS**

Gas Name	Values of UUC	Values on Reference Standard	Error in Reading of UUC	Uncertainty of Measurement
H2S Gas (25 ppm)	25.0	25	0	± 0.2

Calibration results were found to conform as per specified accuracy requirements. Above Instrument has **PASSED** its Calibration.

The results are as found (no adjustment done).

The results are post adjustment.















