

Certificate of Calibration

Al Takamul Yard, North Rumailah South Iraq

• Phone : +964 7810009138 • www.qualitycontrol-iraq.com • E-mail: op@qualitycontrol-iraq.com

Date of Issue: July 3, 2025

REQUEST NUMBER : By Contract	APPROVED BY:  Asst. Manager
JOB NUMBER : QC-CAL-25248	
CERTIFICATE NUMBER : QC-CAL-25248-22	



CUSTOMER DETAILS

Name : **Daqing Drilling Company Iraq**
 Address : Basra, South Iraq
 Department : DQ037

EQUIPMENT IDENTIFICATION AND SPECIFICATIONS

Description : **RCD Tester**
 Manufacturer : UNI-T
 Model : UT-582
 Serial Number : **CALDQ25041718**
 Type : Digital
 Range : 10 to 500 mA
 Calibrated Range : 10 to 500 mA
 Readability : 0.01 mA
 Calibration Date : July 3, 2025
Calibration Due : July 2, 2026
 Last Calibration : NA



ENVIRONMENTAL CONDITIONS DURING TEST

Ambient Temperature : 22.5 ± 1°C Humidity : 40.6 ± 5 % RH Atm. Pressure : 1013.6 ± 10 hPa

CALIBRATION METHOD

The deviations of the measurements obtained from UUC with respect to reference standards are determined to obtain the error.

REFERENCE EQUIPMENT USED :

DESCRIPTION	MAKE	MODEL #	SERIAL #	CALIBRATION DATE	CALIBRATION DUE DATE
Electrical Test Equipment Calibrator	TRANSMILLE	3200A	2806902	8/20/2024	8/19/2025

CALIBRATION TEST RESULTS

Range Function	Selected Values (A)	Measured Values (M)	Deviation (D=M-A)	Expanded Uncertainty (U)
	mA	mA	mA	± mA
x1/2	10.00	10.02	0.02	0.05
x1/2	20.0	20.0	-0.02	0.05
x1	100.00	99.96	-0.04	0.05
x1	300.0	300.0	0.04	0.05
x5	20	20.0	0.02	0.05
x5	30	30.0	0.03	0.05

Calibration results were found to conform as per specified accuracy requirements. Above Instrument has **PASSED** its Calibration.

- The results are as found (no adjustment done).
 The results are post adjustment.



Corporate Partner

Address : North Rumaila, Al Takamul Yard
 Contact : +9647810009138