

Certificate of Calibration

Al Takamul Yard, North Rumailah South Iraq

• Phone : +964 7810009138 • www.qualitycontrol-iraq.com • E-mail: op@qualitycontrol-iraq.com

Date of Issue: July 3, 2025

REQUEST NUMBER	: By Contract	APPROVED BY LAB IN CHARGE
JOB NUMBER	: QC-CAL-25248	Asjad Ra
CERTIFICATE NUMBER	: QC-CAL-25248-23	



CUSTOMER DETAILS

Name : **Daqing Drilling Company Iraq**
 Address : Basra, South Iraq
 Department : DQ037

EQUIPMENT IDENTIFICATION AND SPECIFICATIONS

Description : **Anemometer**
 Manufacturer : MASTECH
 Model : MS6252A
 Location : DQ037
 Serial Number : **CALDQ25041714**
 Type : Digital
 Range : 0.8 m/s to 30 m/s
 Calibrated Range : 0.8 m/s to 30 m/s
 Readability : 0.1 m/s
 Accuracy : ±2% rdg
 Calibrated By : Hussein Alaa
 Calibration Date : July 3, 2025
Calibration Due : July 2, 2026



ENVIRONMENTAL CONDITIONS DURING TEST

Ambient Temperature : 31.0 ± 2°C Humidity : 44.0 ± 5 % RH Atm. Pressure : 1011.0 ± 10 hPa

CALIBRATION METHOD

The above equipment has been calibrated in accordance with ASTM D5096-02

REFERENCE EQUIPMENT USED :

DESCRIPTION	SERIAL NO.	MAKE / MODEL	CALIBRATION DATE	CALIBRATION DUE DATE
Benchtop Wind Tunnel	5668331	OMEGA/ WTM-1000	15 Dec, 2024	14 Dec, 2025

CALIBRATION TEST RESULTS

Measured Values on Unit Under Calibration (M)	Ref. Standard Values (A)	Deviation (D=M-A)	Deviation %
m/s	m/s	m/s	%
5.0	5.00	0.00	0.0
10.0	10.00	0.00	0.0
15.2	15.00	0.20	1.3
20.3	20.00	0.30	1.5
30.3	30.00	0.30	1.0

Results :

The above Value Indicates That the Instrument complies with the Specified Maximum Permissible error tolerance.
 At The Time of Calibration. and ready for use.

- The results are as found (no adjustment done).
 The results are post adjustment.