

Calibration Certificate

Al Takamul Yard North Rumailah, Iraq

• Phone : +964 7810009138 • www.qualitycontrol-iraq.com • E-mail: op@qualitycontrol-iraq.com

Date of Issue: July 3, 2025

Page 1 of 2

REQUEST NUMBER : By Contract	APPROVED BY: CALIB IN CHARGE QC
JOB NUMBER : QC-CAL-25248	
CERTIFICATE NUMBER : QC-CAL-25248-26	



CUSTOMER DETAILS

Name : **Daqing Drilling Company Iraq**
 Department : Basra, South Iraq
 Address : DQ037

EQUIPMENT IDENTIFICATION AND SPECIFICATIONS

Description : **Earth Ground Clamp Meter**
 Type of Indication : Digital
 Manufacturer : UNI-T
 Model : UT276A+
 Serial Number : CALDQ25041717
Calibrated Range :
 Loop Current : 0.02 mA to 30.00 A

 Resolution : 0.001mA, 0.01 mA, 0.1 mA, 0.001 A, 0.01A

 Loop Resistance : 0.025Ω to 1200 Ω

 Resolution : 0.001 Ω, 0.01 Ω, 0.1 Ω, 1 Ω
 Accuracy : ±(1%+0.01) to ±(10%+10)
 As Found : In Tolerance
 Calibrated By : Hussein Alaa
 Calibration Date : July 3, 2025
Calibration Due : **July 2, 2026**
 Last Calibration : NA



ENVIRONMENTAL CONDITIONS DURING TEST

Ambient Temperature : 31 °C ± 2 °C
 Relative Humidity : 20 %RH ± 5 %RH

CALIBRATION METHOD

The above equipment has been calibrated in accordance with QC Calibration Procedure # QC/CP/E/01
 The deviations of the measurements obtained from UUC with respect to reference standards are determined to obtain the error.

TRACEABILITY

The measurements made by Quality Control Labs, realize the physical units of measurements (SI), through its state of the art calibration standards that are controlled and maintained by QC.

REFERENCE EQUIPMENT USED :

DESCRIPTION	MAKE	MODEL #	SERIAL #	CALIBRATION DATE	TRACEABILITY
Multifunction Calibrator	Fluke, USA	5522A	2806902	8/20/2024	8/19/2025
Ground Resistance Tester	Hioki	FT6380	253358	8/20/2024	8/19/2025
Standard Loop Resistance for Clamp On Ground Tester					



CERTIFICATE OF CALIBRATION

CERTIFICATE NUMBER

QC-CAL-25248-26

Page 2 of 2

REQUEST NUMBER: By Mail
JOB NUMBER: QC-CAL-25248

CALIBRATION TEST RESULTS

Measurement Data for Milliampere

Readings on UUC mA	Readings on Ref. Standard mA	Error mA	Uncertainty ± (mA)
0.02	0.20000	-0.180	0.01
2.002	2.00000	0.002	0.01
20.016	20.0000	0.016	0.01
30.024	30.0000	0.024	0.01

Measurement Data for Amperes

Readings on UUC A	Readings on Ref. Standard A	Error A	Uncertainty ± (A)
4.002	4.00000	0.002	0.0
9.992	10.0000	-0.008	0.2
19.998	20.0000	-0.002	0.02

Measurement Data for AC Voltage @ 50 Hz

Readings on UUC Ω	Readings on Ref. Standard Ω	Error Ω	Uncertainty ± (Ω)
0.476	0.474	0.002	0.0
0.503	0.500	0.003	0.001
10.03	10.000	0.030	0.02
101.00	100.00	1.000	0.2

Calibration results were found to conform as per specified accuracy requirements. Above Instrument has **PASSED** its Calibration.

The reported expanded uncertainty is based on a standard uncertainty multiplied by a coverage factor $k=2$, providing a level of confidence of approximately 95%. The uncertainty evaluation has been carried out in accordance with international practice.

DEVIATION FROM STANDARD METHOD : None

REMARK (S) : The results are as found (no adjustment done).
 The results are post adjustment.