

Certificate of Calibration

Al Takamul Yard, North Rumailah South Iraq

• Phone : +964 7810009138 • www.qualitycontrol-iraq.com • E-mail: op@qualitycontrol-iraq.com

Date of Issue: July 5, 2025

Page 1 of 1

REQUEST NUMBER : By Mail	CUSTOMER DETAILS	
JOB NUMBER : QC-CAL-25250	Name : Halliburton Worldwide IRAQ (TSS)	
CERTIFICATE NUMBER : QC-CAL-25250-12	Address : Western Burjesia, Oil Street, District Zubair-South Iraq	

Description : **Pressure Gauge**
 Manufacturer : WIKA
 Model : EN837-1
 Serial Number : **TSS772507**
 Type : Analog
 Location : Separator SAP # 11484041
 Range : 0 to 230 psi
 Calibrated Range : 0 to 230 psi
 Resolution : 10 psi
 Tolerance : ± 1.6 % of Full Scale
 As found : In Tolerance
 Calibration Date : July 5, 2025
Calibration Due : July 4, 2026



DEVIATION FROM STANDARD METHOD : None

REMARK (S) :

1. No adjustments were made prior to calibration.
2. No accessories were fitted.
3. Values were rounded in computation.

ENVIRONMENTAL CONDITIONS DURING TEST

Ambient Temperature : 29.0 ± 2°C Humidity : 44.0 ± 5 % RH Atm. Pressure : 1011.0 ± 10 hPa

CALIBRATION METHOD

The above equipment has been calibrated in accordance with Calibration Procedure QC/CP/R/03 Rev. 01

The above instrument has been calibrated in increasing and decreasing pressure mode using a Digital Pressure Calibrator as the reference standard having traceability to Internationally recognised standards.

REFERENCE EQUIPMENT USED


DESCRIPTION	SERIAL NO.	MAKE / MODEL	CALIBRATION DATE	CALIBRATION DUE DATE
Digital Pressure Gauge	468802	AMETEK / XP2i	5 Nov, 2024	4 Nov, 2025

CALIBRATION TEST RESULTS

Measured Values on Unit Under Calibration (M)	Ref. Standard Values (A)	Deviation (D=M-A)	Full Scale Error %
0	0	0	%
0	0.00	0.00	0.0
60	61.00	-1.00	-0.4
120	121.00	-1.00	-0.4
180	182.00	-2.00	-0.9
230	232.00	-2.00	-0.9

Calibration results were found to conform as per specified accuracy requirements. Above Instrument has **PASSED** its Calibration.

- The results are as found (no adjustment done).
 The results are post adjustment.

CALIBRATED BY  Abdulrahman Loay	REVIEWED & APPROVED BY LAB INCHARGE  LAB INCHARGE Asjad Rafiq	CLIENT
--	---	--------