

Certificate of Calibration

Al Takamul Yard, North Rumailah South Iraq

• Phone : +964 7810009138 • www.qualitycontrol-iraq.com • E-mail: op@qualitycontrol-iraq.com

Date of Issue: July 5, 2025

Page 1 of 1

| | | |
|---|--|--|
| REQUEST NUMBER : By Mail | CUSTOMER DETAILS | |
| JOB NUMBER : QC-CAL-25250 | Name : Halliburton Worldwide IRAQ (TSS) | |
| CERTIFICATE NUMBER : QC-CAL-25250-26 | Address : Western Burjesia, Oil Street, District Zubair-South Iraq | |

Description : **Pressure Gauge**
 Manufacturer : STEWARTS-USA
 Model : N/A
 Serial Number : **HBPB-005**
 Type : Analog
 Location : Pressure Pump
 Range : 0 to 20000 psi
 Calibrated Range : 0 to 15000 psi
 Resolution : 200 psi
 Tolerance : ± 1.0 % of Full Scale
 As found : In Tolerance
 Calibration Date : July 5, 2025
Calibration Due : July 4, 2026



DEVIATION FROM STANDARD METHOD : None

REMARK (S) :

1. No adjustments were made prior to calibration.
2. No accessories were fitted.
3. Values were rounded in computation.

ENVIRONMENTAL CONDITIONS DURING TEST

Ambient Temperature : 29.0 ± 2°C Humidity : 44.0 ± 5 % RH Atm. Pressure 1011.0 ± 10 hPa

CALIBRATION METHOD

The above equipment has been calibrated in accordance with Calibration Procedure QC/CP/R/03 Rev. 01

The above instrument has been calibrated in increasing and decreasing pressure mode using a Digital Pressure Calibrator as the reference standard having traceability to Internationally recognised standards.

REFERENCE EQUIPMENT USED


| DESCRIPTION | SERIAL NO. | MAKE / MODEL | CALIBRATION DATE | CALIBRATION DUE DATE |
|------------------------|------------|---------------|------------------|----------------------|
| Digital Pressure Gauge | 468802 | AMETEK / XP2i | 5 Nov, 2024 | 4 Nov, 2025 |

CALIBRATION TEST RESULTS

| Measured Values on Unit Under Calibration (M) | Ref. Standard Values (A) | Deviation (D=M-A) | Full Scale Error % |
|---|--------------------------|-------------------|--------------------|
| 0 | 0 | 0 | % |
| 0 | 0.00 | 0.00 | 0.0 |
| 4000 | 3985.00 | 15.00 | 0.1 |
| 8000 | 7981.00 | 19.00 | 0.1 |
| 12000 | 11974.00 | 26.00 | 0.1 |
| 15000 | 14970.00 | 30.00 | 0.2 |

Calibration results were found to conform as per specified accuracy requirements. Above Instrument has **PASSED** its Calibration.

- The results are as found (no adjustment done).
 The results are post adjustment.

| | | |
|--|--|-----------------------|
| CALIBRATED BY  Abdulrahman Loay | REVIEWED & APPROVED BY LAB INCHARGE  LAB INCHARGE | CLIENT Asjad Rafiq |
|--|--|-----------------------|