

## Certificate of Calibration

Al Takamul Yard, North Rumailah South Iraq

• Phone : +964 7810009138 • www.qualitycontrol-iraq.com • E-mail: op@qualitycontrol-iraq.com

Date of Issue: July 5, 2025

Page 1 of 1

|   |  |  |
|---|--|--|
| REQUEST NUMBER : By Mail                    | CUSTOMER DETAILS   |  |
| JOB NUMBER : QC-CAL-25250                   | Name : <b>Halliburton Worldwide IRAQ (TSS)</b>                     |  |
| CERTIFICATE NUMBER : <b>QC-CAL-25250-34</b> | Address : Western Burjesia, Oil Street, District Zubair-South Iraq |  |

Description : **Pressure Gauge**  
 Manufacturer : WIKA  
 Model : 2302211107  
 Serial Number : **TSS772510**  
 Type : Analog  
 Location : TSS Yard  
 Range : 0 to 1000 psi  
 Calibrated Range : 0 to 1000 psi  
 Resolution : 50 psi  
 Tolerance : ± 1.0 % of Full Scale  
 As found : In Tolerance  
 Calibration Date : July 5, 2025  
**Calibration Due : July 4, 2026**



**DEVIATION FROM STANDARD METHOD : None**

**REMARK (S) :**

1. No adjustments were made prior to calibration.
2. No accessories were fitted.
3. Values were rounded in computation.

### ENVIRONMENTAL CONDITIONS DURING TEST

Ambient Temperature : 29.0 ± 2°C Humidity : 44.0 ± 5 % RH Atm. Pressure 1011.0 ± 10 hPa

### CALIBRATION METHOD

The above equipment has been calibrated in accordance with Calibration Procedure QC/CP/R/03 Rev. 01

The above instrument has been calibrated in increasing and decreasing pressure mode using a Digital Pressure Calibrator as the reference standard having traceability to Internationally recognised standards.

### REFERENCE EQUIPMENT USED


| DESCRIPTION            | SERIAL NO. | MAKE / MODEL  | CALIBRATION DATE | CALIBRATION DUE DATE |
|------------------------|------------|---------------|------------------|----------------------|
| Digital Pressure Gauge | 468802     | AMETEK / XP2i | 5 Nov, 2024      | 4 Nov, 2025          |

### CALIBRATION TEST RESULTS

| Measured Values on Unit Under Calibration (M) | Ref. Standard Values (A) | Deviation (D=M-A) | Full Scale Error % |
|---|--------------------------|-------------------|--------------------|
| 0   | 0                        | 0                 | %                  |
| 0   | 0.00                     | 0.00              | 0.0                |
| 250   | 252.00                   | -2.00             | -0.2               |
| 500   | 503.00                   | -3.00             | -0.3               |
| 750   | 755.00                   | -5.00             | -0.5               |
| 1000  | 1006.00                  | -6.00             | -0.6               |

Calibration results were found to conform as per specified accuracy requirements. Above Instrument has **PASSED** its Calibration.

- The results are as found (no adjustment done).  
 The results are post adjustment.

|  |  |        |
|--|--|--------|
| CALIBRATED BY<br><br>Abdulrahman Loay | REVIEWED & APPROVED BY LAB INCHARGE<br><br>LAB INCHARGE Asjad Rafiq | CLIENT |
|--|--|--------|