

WITNESSING PRESSURE TEST Certificate

Client:	SLB Basra	Report No:	QC-SLB-25235-14
Work Location:	(Testing Services)	Test Date:	14-Sep-2025
Reference:	SLB Procedure - MFG Specification	Next Test Date:	13-Sep-2026
Inspection Method :	√ VT √ Witness		

FUNCTIONAL TEST

SLB Carried out the test according to their specification and this was witnessed by Quality Control – Iraq.

The following test steps that performed on the centrifuge:

- The sample was heated for a few minutes in the heating pocket.
- Then put inside the centrifuge and rotation started.
- The result shown in the tube read was in three phases (water, oil, and sediments).

Test Sample:	Crude Oil	Duration:	10 Minutes
---------------------	-----------	------------------	------------

Result: BSW 3.0%

Serial No	Description	
RS15K11648	Walker Centrifuge	

Result:	The functional test carried out by SLB, witnessed by Quality Control Iraq, and found to be working correctly as per the above steps at the time of test.
----------------	--

Customer:		Witnessed By:	M-SHAHZAD
Signature:		Signature:	

