


## Certificate of Calibration

Al Takamul Yard, North Rumailah South Iraq

• Phone : +964 7810009138 • www.qualitycontrol-iraq.com • E-mail: op@qualitycontrol-iraq.com

Date of Issue: March 5, 2026

Page 1 of 1

REQUEST NUMBER : By Contract	APPROVED BY LAB INCHARGE QC
JOB NUMBER : QC-CAL-26076	
CERTIFICATE NUMBER : <b>QC-CAL-26076-01</b>	

### CUSTOMER DETAILS

Name : **Egyptian Maintenance Company(EMC) IRAQ**  
 Address : Iraq, Basra

### EQUIPMENT IDENTIFICATION AND SPECIFICATIONS

Description : **Push off Adhesion Tester**  
 Manufacturer : Elcometer  
 Model : F506-D  
 Serial Number : NL23088  
 Type : Analog  
 Range : 300 to 3600 psi  
 Tolerance : ±1.5% ( as Per Manufacturer )  
 As found : Good  
 Calibration Date : March 5, 2026  
**Calibration Due : March 4, 2027**



### ENVIRONMENTAL CONDITIONS DURING TEST

Ambient Temperature : 22.0 ± 2°C Humidity : 40.0 ± 5 % RH Atm. Pressure 1012.0 ± 10 hPa

### CALIBRATION METHOD

The above equipment has been calibrated in accordance with Calibration Procedure QC/CP/R/06 Rev. 01

### REFERENCE EQUIPMENT USED

DESCRIPTION	SERIAL NO.	MAKE / MODEL	CALIBRATION DATE	CALIBRATION DUE DATE
Push off Adhesion Tester	NL58390	Elcometer / 508	November 3, 2025	November 2, 2026

### CALIBRATION TEST RESULTS

the Unit Under Calibration and the Reference Equipment, all use same surfaces During Do the Test and at the Same Temperature.

Measured Values on Unit Under Calibration (M)	Ref. Standard Values (A)	Deviation (D=M-A)	Tolerance Limit (±1.5%)
psi	psi	psi	%
960	952.00	8.0	0.83

#### Results:

Calibration results were found to conform as per specified accuracy requirements. Above Instrument has **PASSED** its Calibration.